

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

METALS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1081

MATRIX: Water

METHOD: 6010D

| | NA | NO | YES |
|--|----|----|-----|
| 1. Calibration Summary met criteria. | | | ✓ |
| 2. ICP Interference Check Sample Results Summary Submitted. | | | ✓ |
| 3. Serial Dilution Summary (if applicable) Submitted. | | | ✓ |
| 4. Laboratory Control Sample Summary (if applicable) Submitted. | | | ✓ |
| 5. Blank Contamination - If yes, list compounds and concentrations in each blank: | | ✓ | |
| 6. Matrix Spike/Matrix Spike Duplicate Recoveries Met Criteria | | | ✓ |
| If not met, list those compounds and their recoveries which fall outside the acceptable range. | | | |
| 7. Sample Duplicate Analysis Met QC Criteria | | | ✓ |
| If not met, list those compounds and their recoveries which fall outside the acceptable range. | | | |
| 8. Digestion Holding Time Met | | | ✓ |
| If not met, list number of days exceeded for each sample: | | | |
| 9. Analysis Holding Time Met | | | ✓ |
| If not met, list those compounds and their recoveries which fall outside the acceptable range. | | | |

ADDITIONAL COMMENTS: The fax and hardcopy is not matching for Sample Q1081-04 due to at the time fax sample analyzed without QC set, but at the time of second review lab noticed QC set was not analyzed therefore this sample analyzed with QC set and reported. Hard copy is reported corrected.

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).

QA REVIEW

Date