

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1081

MATRIX: Water

METHOD: 8270-Modified/3510

| | | NA | NO | YES |
|----|---|----|----|-----|
| 1. | Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks) | | | ✓ |
| 2. | GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ) | | | ✓ |
| 3. | GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series. | | | ✓ |
| 4. | GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series. | | | ✓ |
| 5. | GC/MS Calibration Requirements. The Initial Calibration met the requirements . The Continuous Calibration File ID BN035926.D met the requirements except for 2,4,6-Tribromophenol, The failure compound not associated with the client parameters list, therefore no corrective action was taken. | | ✓ | |
| 6. | Blank Contamination - If yes, list compounds and concentrations in each blank: | | ✓ | |
| 7. | Surrogate Recoveries Meet Criteria | | ✓ | |

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

The Surrogate recoveries met the acceptable criteria except for PB166053BL [2-Fluorobiphenyl - 127%, Nitrobenzene-d5 - 138%], PB166053BS [Nitrobenzene-d5 - 118%], PB166053BSD [Nitrobenzene-d5 - 113%], RW7-SP100-20250113 [2-Fluorobiphenyl - 117%, Nitrobenzene-d5 - 120%, Terphenyl-d14 - 145%], RW7-SP100-20250113DL [Nitrobenzene-d5 - 122%, Terphenyl-d14 - 138%], RW7-SP201-20250113 [2-Fluorobiphenyl - 108%, Nitrobenzene-d5 - 114%, Terphenyl-d14 - 135%], RW7-SP302-20250113 [2-Fluorobiphenyl - 108%, Nitrobenzene-d5 - 123%, Terphenyl-d14 - 138%], RW7-SP303-20250113 [2-Fluorobiphenyl - 123%, Nitrobenzene-d5 - 128% and Terphenyl-d14 - 152%], The failure surrogates not associated with the client parameters list, therefore no corrective action was taken.

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

(CONTINUED)

| | | NA | NO | YES |
|-----|--|----|----|-----|
| 8. | Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria | | | ✓ |
| | If not met, list those compounds and their recoveries which fall outside the acceptable range. | | | |
| | The Blank Spike met requirements for all samples . | | | |
| | The Blank Spike Duplicate met requirements for all samples . | | | |
| 9. | Internal Standard Area/Retention Time Shift Meet Criteria | | | ✓ |
| | Comments: | | | |
| 10. | Extraction Holding Time Met | | | ✓ |
| | If not met, list number of days exceeded for each sample: | | | |
| 11. | Analysis Holding Time Met | | | ✓ |
| | If not met, list number of days exceeded for each sample: | | | |

ADDITIONAL COMMENTS:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is)."

Sample RW7-SP100-20250113 was diluted due to high concentration.

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the % RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date