



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1081
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1081 **SAS No.:** Q1081
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Iron	86.4	100	86	40 - 160	P	01/16/2025	19:36	LB134315
CRI01	Iron	97.7	100	98	40 - 160	P	01/17/2025	11:35	LB134334