

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1081 SAS No.: Q1081 SDG No.: Q1081
Instrument ID: BNA_N Calibration Date(s): 01/02/2025 01/02/2025
Calibration Time(s): 11:28 15:04

LAB FILE ID:		RRF0.1 = BN035871.D		RRF0.2 = BN035872.D		RRF0.4 = BN035873.D		
		RRF0.8 = BN035874.D		RRF1.6 = BN035875.D		RRF3.2 = BN035876.D		
COMPOUND	RRF0.1	RRF0.2	RRF0.4	RRF0.8	RRF1.6	RRF3.2	RRF	% RSD
2-Methylnaphthalene-d10	0.547	0.536	0.527	0.519	0.556	0.524	0.536	2.5
Fluoranthene-d10	0.988	0.952	0.978	0.959	1.028	1.010	0.993	3.3
2-Fluorophenol	1.031	1.009	0.952	0.958	0.997	0.956	0.981	3.1
Phenol-d6	1.351	1.255	1.180	1.197	1.215	1.163	1.219	5.4
Nitrobenzene-d5	0.346	0.307	0.296	0.302	0.319	0.317	0.317	5.5
2-Fluorobiphenyl	1.776	1.675	1.708	1.765	1.823	1.779	1.755	2.8
2,4,6-Tribromophenol	0.164	0.165	0.189	0.189	0.207	0.211	0.192	11.4
Terphenyl-d14	0.814	0.813	0.790	0.777	0.828	0.776	0.797	2.6
1,4-Dioxane	0.454	0.422	0.373	0.388	0.391	0.377	0.397	7.5

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1