

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011525\
 Data File : BN035935.D
 Acq On : 15 Jan 2025 17:36
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 15 18:18:53 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	69	0.00
2	1,4-Dioxane	0.397	0.388	2.3	72	0.00
3	n-Nitrosodimethylamine	0.693	0.853	-23.1	88	-0.01
4 S	2-Fluorophenol	0.981	0.914	6.8	67	-0.01
5 S	Phenol-d6	1.219	1.108	9.1	65	0.00
6	bis(2-Chloroethyl)ether	0.929	0.832	10.4	62	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	63	0.00
8 S	Nitrobenzene-d5	0.317	0.340	-7.3	73	-0.01
9	Naphthalene	1.123	1.110	1.2	65	-0.01
10	Hexachlorobutadiene	0.365	0.393	-7.7	71	0.00
11 SURR	2-Methylnaphthalene-d10	0.536	0.523	2.4	63	0.00
12	2-Methylnaphthalene	0.695	0.693	0.3	64	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	62	0.00
14 S	2,4,6-Tribromophenol	0.192	0.267	-39.1#	88	-0.01
15 S	2-Fluorobiphenyl	1.755	1.796	-2.3	65	0.00
16	Acenaphthylene	1.884	1.865	1.0	64	0.00
17	Acenaphthene	1.235	1.304	-5.6	67	0.00
18	Fluorene	1.359	1.352	0.5	64	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	63	-0.01
20	4,6-Dinitro-2-methylphenol	0.069	0.101	-46.4#	95	0.00
21	4-Bromophenyl-phenylether	0.274	0.285	-4.0	67	-0.01
22	Hexachlorobenzene	0.374	0.378	-1.1	67	0.00
23	Atrazine	0.184	0.211	-14.7	75	0.00
24	Pentachlorophenol	0.132	0.180	-36.4#	95	0.00
25	Phenanthrene	1.167	1.132	3.0	62	-0.01
26	Anthracene	1.062	1.029	3.1	64	0.00
27 SURR	Fluoranthene-d10	0.993	0.980	1.3	63	0.00
28	Fluoranthene	1.361	1.334	2.0	63	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	61	0.00
30	Pyrene	1.624	1.606	1.1	63	0.00
31 S	Terphenyl-d14	0.797	0.817	-2.5	63	0.00
32	Benzo(a)anthracene	1.410	1.420	-0.7	65	-0.02
33	Chrysene	1.475	1.442	2.2	61	0.00
34	Bis(2-ethylhexyl)phthalate	0.574	0.715	-24.6	75	0.00
35 I	Perylene-d12	1.000	1.000	0.0	63	-0.01
36	Indeno(1,2,3-cd)pyrene	1.577	1.540	2.3	68	-0.02
37	Benzo(b)fluoranthene	1.374	1.316	4.2	64	-0.01
38	Benzo(k)fluoranthene	1.360	1.375	-1.1	69	-0.01
39 C	Benzo(a)pyrene	1.190	1.149	3.4	66	-0.01
40	Dibenzo(a,h)anthracene	1.255	1.236	1.5	70	-0.02
41	Benzo(g,h,i)perylene	1.403	1.319	6.0	66	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4