

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/13/25
Project:	RW7B - 112G08005-WE13		Date Received:	01/14/25
Client Sample ID:	RW7-SP201-20250113		SDG No.:	Q1081
Lab Sample ID:	Q1081-02		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	990	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN035929.D	1	01/14/25 11:00	01/15/25 12:15	PB166053

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.44		30 - 150		110%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.52		30 - 150		131%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.46	*	55 - 111		114%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.43	*	53 - 106		108%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.54	*	58 - 132		135%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	707		7.832			
1146-65-2	Naphthalene-d8	1340		10.622			
15067-26-2	Acenaphthene-d10	689		14.458			
1517-22-2	Phenanthrene-d10	1430		17.194			
1719-03-5	Chrysene-d12	1320		21.376			
1520-96-3	Perylene-d12	1480		23.674			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products