

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL01

Lab Code: CHEM Case No.: Q1083 SAS No.: Q1083 SDG No.: Q1083

Instrument ID: MSVOA_X Calibration Date/Time: 01/16/2025 09:02

Lab File ID: VX044660.D Init. Calib. Date(s): 01/16/2025 01/16/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:39 10:12

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.665	1.726		3.66	
Toluene	1.975	1.972		-0.15	
Ethyl Benzene	2.156	2.181		1.16	
m/p-Xylenes	0.792	0.817		3.16	
o-Xylene	0.795	0.823		3.52	
1,2-Dichloroethane-d4	2.896	2.867		-1	
Toluene-d8	1.632	1.620		-0.74	
4-Bromofluorobenzene	0.583	0.578		-0.86	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.