

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1083

Client: Tully Environmental, Inc

Analytical Method: SW624.1

Datafile : VX044667.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0116WBSD01	Benzene	20	20.9	ug/L	104	0		65	135	20
	Toluene	20	20.3	ug/L	102	4		70	130	20
	Ethyl Benzene	20	20.2	ug/L	101	5		60	140	20
	m/p-Xylenes	40	41.7	ug/L	104	5		87	111	20
	o-Xylene	20	20.9	ug/L	104	5		87	111	20