

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL01
 Lab Code: CHEM Case No.: Q1083 SAS No.: Q1083 SDG NO.: Q1083
 Lab File ID: VX044660.D Date Analyzed: 01/16/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:02
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	33434	4.89	183316	6.76	165841	10.05
UPPER LIMIT	66868	5.391	366632	7.257	331682	10.549
LOWER LIMIT	16717	4.391	91658	6.257	82920.5	9.549
EPA SAMPLE NO.						
001-WILLETTS-PT-BLVD (DEC)	34454	4.89	193425	6.76	171637	10.05
002-35TH-AVE (DEC)	31297	4.89	176011	6.76	149063	10.06
VX0116WBL01	31761	4.89	169032	6.76	146905	10.06
VX0116WBS01	28646	4.89	161068	6.75	140626	10.05
VX0116WBSD01	29143	4.89	154664	6.75	141140	10.05

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.