

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51943 MA No.: SDG No.: MC0B99
Instrument ID: P5 Start Date: 01/20/2025
Analytical Method: ICP-AES Run Batch: LB134352
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Calcium	0	-0.51	0	5000	5200	4	62500	65000	4
Magnesium	0	-0.35	0	5000	5100	2	62500	65000	4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51943 MA No.: SDG No.: MC0B99
Instrument ID: P5 Start Date: 01/20/2025
Analytical Method: ICP-AES Run Batch: LB134352
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Calcium	125000	130000	4	250000	250000	0	500000	490000	-2
Magnesium	125000	130000	4	250000	250000	0	500000	490000	-2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51943 MA No.: SDG No.: MC0B99
Instrument ID: P5 Start Date: 01/20/2025
Analytical Method: ICP-AES Run Batch: LB134352
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Calcium	800000	800000	0						
Magnesium	800000	800000	0						