

FORM 4 - IN
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51943 MA No. : SDG No.: MC0B99
Analytical Method: ICP-AES ICSA Source: EPA-1211
Instrument ID: P5 ICSB Source: EPA-0710
Run Batch: LB134352
Concentration Units: ug/L

Analyte	True		Found			
	ICSA	ICSAB	ICSA	%R	ICSAB	%R
Aluminum	255000	247000	240000	95	250000	101
Calcium	245000	235000	250000	101	260000	109
Iron	101000	99300	100000	101	110000	106
Magnesium	255000	248000	250000	98	260000	104
Potassium	0	0	-1200		-2400	
Sodium	0	0	2000		2200	