

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51943 MA No. : SDG No.: MC0B99
Initial Calibration Verification Source : MP84083
Continuing Calibration Verification Source : MP84085
Run Batch: LB134306 Analytical Method: CVAA
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV058				ID: CCV039				ID: CCV040		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	3.7	94		5.0	4.8	97		5.7	114	