EPA SAMPLE NO.

MC0BA2S

FORM 5A - IN MATRIX SPIKE SAMPLE RECOVERY

Lab Name:	Chemtech Cor	nsulting Grou	ıp	Contract:	EPW14030				
Lab Code:	ACE	Case No.:	51943	MA No.:		SDG No.:	MC0B99		
Matrix :	Water			Analytical	Method:	CVAA			
% Solids:									
Concentrat	ion Units (μg	/L, mg/L or n	ng/kg dry weigh	nt): ug/1	L				

Analyte	Control Limit %R	Spiked Sample Result (SSR)	Q	Sample Result (SR)	Q	Spike Added (SA)	%R	Q
Mercury	75 - 125	1.1		0.2	U	1.0	114	