


LB134397

SFAM01.1

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Std
0.00	0.000	0.062	0.062	187	0.000	187					
0.05	0.050										
0.20	0.200	0.233	0.033	949	0.0 %	949					
2.50	2.500	2.500	0.000	11047	0.0 %	11047					
5.00	5.000	4.819	-0.181	21374	0.0 %	21374					
7.50	7.500	7.503	0.003	33332	0.0 %	33332					
10.0	10.000	10.083	0.083	44824	0.0 %	44824					

LB134397
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	187	-		0 SFAM01.1	PPB	1/24/2025 13:17	Std
	0.2 S01	949	-		0.2 SFAM01.1	PPB	1/24/2025 13:20	Std
	2.5 S02	11047	-		2.5 SFAM01.1	PPB	1/24/2025 13:22	Std
	5 S03	21374	-		5 SFAM01.1	PPB	1/24/2025 13:24	Std
	7.5 S04	33332	-		7.5 SFAM01.1	PPB	1/24/2025 13:27	Std
	10 S05	44824	-		10 SFAM01.1	PPB	1/24/2025 13:29	Std
ICV068	ICV068	16057	3.6249 -		SFAM01.1	PPB	1/24/2025 13:32	SMPL
ICB068	ICB068	89	0.04 -		SFAM01.1	PPB	1/24/2025 13:34	SMPL
CCV068	CCV068	21601	4.8695 -		SFAM01.1	PPB	1/24/2025 13:36	SMPL
CCB068	CCB068	85	0.0391 -		SFAM01.1	PPB	1/24/2025 13:39	SMPL
PB166234BL	PBW234	249	0.0759 -		SFAM01.1	PPB	1/24/2025 13:41	SMPL
Q1087-01	MCOAF9	224	0.0703 -		SFAM01.1	PPB	1/24/2025 13:43	SMPL
Q1087-02	MCOAH2	260	0.0784 -		SFAM01.1	PPB	1/24/2025 13:45	SMPL
Q1087-03	MCOAH3	103	0.0431 -		SFAM01.1	PPB	1/24/2025 13:48	SMPL
Q1087-04	MCOAF8	35	0.0278 -		SFAM01.1	PPB	1/24/2025 13:50	SMPL
Q1087-05	MCOAF8D	323	0.0925 -		SFAM01.1	PPB	1/24/2025 13:52	SMPL
Q1087-06	MCOAF8S	4464	1.0222 -		SFAM01.1	PPB	1/24/2025 13:55	SMPL
Q1087-07	MCOAE9	932	0.2292 -		SFAM01.1	PPB	1/24/2025 13:57	SMPL
Q1087-08	MCOAG4	182	0.0608 -		SFAM01.1	PPB	1/24/2025 13:59	SMPL
Q1087-09	MCOAH0	947	0.2326 -		SFAM01.1	PPB	1/24/2025 14:01	SMPL
Q1087-10	MCOAE6	151	0.0539 -		SFAM01.1	PPB	1/24/2025 14:04	SMPL
Q1087-11	MCOAF0	319	0.0916 -		SFAM01.1	PPB	1/24/2025 14:06	SMPL
Q1087-12	MCOAF1	173	0.0588 -		SFAM01.1	PPB	1/24/2025 14:08	SMPL
Q1087-13	MCOAF5	2013	0.4719 -		SFAM01.1	PPB	1/24/2025 14:10	SMPL
Q1087-14	MCOAF6	-68	0.0047 -		SFAM01.1	PPB	1/24/2025 14:13	SMPL
Q1087-15	MCOAF7	104	0.0433 -		SFAM01.1	PPB	1/24/2025 14:15	SMPL
Q1087-16	MCOAE3	89	0.04 -		SFAM01.1	PPB	1/24/2025 14:17	SMPL
Q1087-17	MCOAE4	295	0.0862 -		SFAM01.1	PPB	1/24/2025 14:19	SMPL
Q1087-18	MCOAE5	167	0.0575 -		SFAM01.1	PPB	1/24/2025 14:22	SMPL
Q1087-19	MCOAF2	605	0.1558 -		SFAM01.1	PPB	1/24/2025 14:24	SMPL
CCV069	CCV069	20870	4.7054 -		SFAM01.1	PPB	1/24/2025 14:26	SMPL
CCB069	CCB069	93	0.0409 -		SFAM01.1	PPB	1/24/2025 14:29	SMPL
Q1087-20	MCOAF3	172	0.0586 -		SFAM01.1	PPB	1/24/2025 14:31	SMPL
Q1087-21	MCOAF4	137	0.0507 -		SFAM01.1	PPB	1/24/2025 14:33	SMPL
Q1087-23	MCOAD7	35158	7.9131 -		SFAM01.1	PPB	1/24/2025 14:38	SMPL
CCV070	CCV070	21020	4.7391 -		SFAM01.1	PPB	1/24/2025 14:45	SMPL
CCB070	CCB070	-7	0.0184 -		SFAM01.1	PPB	1/24/2025 14:47	SMPL