

FORM 13-IN
ICP-MS TUNE

EPA SAMPLE NO.

TUNE

Lab Name: Alliance Technical Group, L Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51953 MA No.: _____ SDG No.: MC0AF9
 Instrument ID: P7 Date: 01/21/2025
 Run Batch: LB134565

Element - Mass	Avg. Measured Mass (u)	Average Peak Width (u)	%Height	%RSD
Be - 9	9.1	0.7	5	0.53
Co - 59	59.0	0.8	5	0.17
In - 113	113.0	0.7	5	0.19
In - 115	115.0	0.7	5	0.89
Mg - 24	24.0	0.8	5	0.70
Mg - 25	25.0	0.7	5	0.62
Mg - 26	26.0	0.8	5	0.70
Pb - 206	206.1	0.8	5	0.30
Pb - 207	207.0	0.8	5	0.18
Pb - 208	208.0	0.8	5	0.40

FORM 13-IN
ICP-MS TUNE

EPA SAMPLE NO.

TUNE

Lab Name: Alliance Technical Group, L Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51953 MA No.: _____ SDG No.: MC0AF9
 Instrument ID: P7 Date: 02/04/2025
 Run Batch: LB134593

Element - Mass	Avg. Measured Mass (u)	Average Peak Width (u)	%Height	%RSD
Be - 9	9.2	0.8	5	1.1
Co - 59	59.0	0.8	5	0.72
In - 113	113.0	0.7	5	0.97
In - 115	115.0	0.8	5	0.79
Mg - 24	24.0	0.8	5	1.4
Mg - 25	25.0	0.8	5	1.1
Mg - 26	26.0	0.8	5	0.92
Pb - 206	206.0	0.8	5	0.90
Pb - 207	206.9	0.7	5	0.84
Pb - 208	207.9	0.7	5	0.64