

FORM 3 - IN
BLANKS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51953 MA No. : _____ SDG No.: MC0AF9
 Preparation Blank Matrix : Water
 Preparation Blank Concentration Units (µg/L, mg/L, mg/kg dry weight, or µg): ug/L
 Analytical Method: ICP-MS Preparation Batch: PB166136
 Run Batch: LB134565 Preparation Method: 200.8

Analyte	Initial Calibration Blank (ug/L)		Continuing Calibration Blank (ug/L)						Preparation Blank/Leachate Extraction Blank	
	ID: ICB001	Q	ID: CCB001	Q	ID: CCB002	Q	ID: CCB003	Q	ID: PBW136	Q
Aluminum	20.0	U	20.0	U	20.0	U	20.0	U	20.0	U
Antimony	2.0	U	0.19	J	0.18	J	2.0	U	2.0	U
Arsenic	1.0	U	1.0	U	1.0	U	1.0	U	1.0	U
Barium	10.0	U	10.0	U	10.0	U	10.0	U	10.0	U
Beryllium	1.0	U	1.0	U	1.0	U	1.0	U	1.0	U
Cadmium	1.0	U	1.0	U	1.0	U	1.0	U	1.0	U
Calcium	500	U	36	J	37	J	500	U	500	U
Chromium	-0.24	J	-0.21	J	-0.22	J	0.21	J	0.21	J
Cobalt	1.0	U	1.0	U	1.0	U	1.0	U	1.0	U
Copper	2.0	U	2.0	U	2.0	U	2.0	U	2.0	U
Iron	200	U	200	U	200	U	200	U	200	U
Lead	1.0	U	0.35	J	0.34	J	1.0	U	1.0	U
Magnesium	500	U	500	U	500	U	500	U	500	U
Manganese	1.0	U	1.0	U	1.0	U	1.0	U	1.0	U
Nickel	1.0	U	1.0	U	1.0	U	1.0	U	1.0	U
Potassium	500	U	500	U	500	U	500	U	500	U
Selenium	5.0	U	5.0	U	5.0	U	5.0	U	5.0	U
Silver	1.0	U	0.1	J	0.09	J	1.0	U	1.0	U
Sodium	500	U	500	U	500	U	500	U	500	U
Thallium	1.0	U	0.1	J	0.08	J	1.0	U	1.0	U
Vanadium	5.0	U	5.0	U	5.0	U	5.0	U	5.0	U
Zinc	5.0	U	5.0	U	5.0	U	5.0	U	5.0	U

FORM 3 - IN
BLANKS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51953 MA No. : _____ SDG No.: MCOAF9
 Preparation Blank Matrix : _____
 Preparation Blank Concentration Units (µg/L, mg/L, mg/kg dry weight, or µg): _____
 Analytical Method: ICP-MS Preparation Batch: _____
 Run Batch: LB134565 Preparation Method: _____

Analyte	Initial Calibration Blank (ug/L)		Continuing Calibration Blank (ug/L)						Preparation Blank/Leachate Extraction Blank	
	ID:	Q	ID: CCB004	Q	ID: CCB005	Q	ID:	Q	ID:	Q
Aluminum			20.0	U	20.0	U				
Antimony			0.25	J	0.18	J				
Arsenic			1.0	U	1.0	U				
Barium			0.56	J	10.0	U				
Beryllium			1.0	U	1.0	U				
Cadmium			1.0	U	1.0	U				
Calcium			54	J	23	J				
Chromium			2.0	U	0.22	J				
Cobalt			1.0	U	1.0	U				
Copper			2.0	U	2.0	U				
Iron			200	U	200	U				
Lead			0.61	J	0.3	J				
Magnesium			500	U	500	U				
Manganese			0.33	J	1.0	U				
Nickel			1.0	U	1.0	U				
Potassium			500	U	500	U				
Selenium			5.0	U	5.0	U				
Silver			0.15	J	1.0	U				
Sodium			500	U	500	U				
Thallium			0.13	J	1.0	U				
Vanadium			5.0	U	5.0	U				
Zinc			5.0	U	5.0	U				

FORM 3 - IN
BLANKS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51953 MA No. : _____ SDG No.: MCOAF9
 Preparation Blank Matrix : _____
 Preparation Blank Concentration Units (µg/L, mg/L, mg/kg dry weight, or µg): _____
 Analytical Method: ICP-MS Preparation Batch: _____
 Run Batch: LB134593 Preparation Method: _____

Analyte	Initial Calibration Blank (ug/L)		Continuing Calibration Blank (ug/L)						Preparation Blank/Leachate Extraction Blank	
	ID: ICB004	Q	ID: CCB015	Q	ID: CCB016	Q	ID: CCB017	Q	ID:	Q
Aluminum	20.0	U	20.0	U	20.0	U	20.0	U		
Antimony	2.0	U	2.0	U	2.0	U	2.0	U		
Arsenic	1.0	U	1.0	U	1.0	U	1.0	U		
Barium	10.0	U	10.0	U	10.0	U	10.0	U		
Beryllium	1.0	U	1.0	U	1.0	U	1.0	U		
Cadmium	1.0	U	1.0	U	1.0	U	1.0	U		
Calcium	500	U	500	U	18	J	500	U		
Chromium	2.0	U	2.0	U	2.0	U	2.0	U		
Cobalt	1.0	U	1.0	U	1.0	U	1.0	U		
Copper	2.0	U	2.0	U	0.5	J	2.0	U		
Iron	200	U	200	U	200	U	200	U		
Lead	1.0	U	1.0	U	0.22	J	0.15	J		
Magnesium	500	U	500	U	24	J	500	U		
Manganese	1.0	U	1.0	U	0.59	J	1.0	U		
Nickel	1.0	U	1.0	U	1.0	U	1.0	U		
Potassium	500	U	500	U	500	U	500	U		
Selenium	5.0	U	5.0	U	5.0	U	5.0	U		
Silver	1.0	U	1.0	U	1.0	U	1.0	U		
Sodium	500	U	500	U	500	U	500	U		
Thallium	1.0	U	1.0	U	1.0	U	1.0	U		
Vanadium	5.0	U	5.0	U	0.04	J	5.0	U		
Zinc	5.0	U	5.0	U	5.0	U	5.0	U		

FORM 3 - IN
BLANKS

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51953 MA No. : _____ SDG No.: MCOAF9
 Preparation Blank Matrix : _____
 Preparation Blank Concentration Units ($\mu\text{g/L}$, mg/L , mg/kg dry weight, or μg): _____
 Analytical Method: ICP-MS Preparation Batch: _____
 Run Batch: LB134593 Preparation Method: _____

Analyte	Initial Calibration Blank ($\mu\text{g/L}$)		Continuing Calibration Blank ($\mu\text{g/L}$)						Preparation Blank/Leachate Extraction Blank	
	ID:	Q	ID: CCB018	Q	ID:	Q	ID:	Q	ID:	Q
Aluminum			20.0	U						
Antimony			2.0	U						
Arsenic			1.0	U						
Barium			10.0	U						
Beryllium			1.0	U						
Cadmium			1.0	U						
Calcium			500	U						
Chromium			2.0	U						
Cobalt			1.0	U						
Copper			2.0	U						
Iron			200	U						
Lead			1.0	U						
Magnesium			500	U						
Manganese			1.0	U						
Nickel			1.0	U						
Potassium			500	U						
Selenium			5.0	U						
Silver			1.0	U						
Sodium			500	U						
Thallium			1.0	U						
Vanadium			5.0	U						
Zinc			5.0	U						