

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51948 MA No.: _____ SDG No.: MC0B00
Instrument ID: CV1 Start Date: 01/31/2025
Analytical Method: CVAA Run Batch: LB134496
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.0080	0	0.2	0.25	23	2.5	2.5	-2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51948 MA No.: SDG No.: MC0B00
Instrument ID: CV1 Start Date: 01/31/2025
Analytical Method: CVAA Run Batch: LB134496
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	4.9	-3	7.5	7.8	4	10	9.9	-1