

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51948 MA No. : SDG No.: MC0B00
 Initial Calibration Verification Source : MP84281
 Continuing Calibration Verification Source : MP84283
 Run Batch: LB134496 Analytical Method: CVAA
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV076				ID: CCV090				ID: CCV091		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	3.7	93		5.0	4.9	98		4.7	94	