

Duplicate Sample Summary

Client: Parkway Generation Operating LLC	SDG No.: Q1097
Project: Linden Generating Station - 402	Sample ID: LB134303BS
Client ID: LB134303BSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TPH	mg/L	+/-18	16.7		17.7		1	5.81		01/16/2025

Duplicate Sample Summary

Client: Parkway Generation Operating LLC	SDG No.: Q1097
Project: Linden Generating Station - 402	Sample ID: Q1097-01
Client ID: 402DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Field pH	pH	+/-20	7.20		7.21		1	0.14		01/15/2025