

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q1097

Project: Linden Generating Station - 402

Run No.: LB134303

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134303BS							
TPH	mg/L	20.0	16.7		84	1	78-114	01/16/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q1097

Project: Linden Generating Station - 402

Run No.: LB134303

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134303BSD							
TPH	mg/L	20.0	17.7		88	1	78-114	01/16/2025