

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Parkway Generation Operating LLC **SDG No.:** Q1097
Contract: PSEG04 **Lab Code:** CHEM **Case No.:** Q1097 **SAS No.:** Q1097
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	517	510	101	95 - 105	P	01/17/2025	11:21	LB134334
	Nickel	508	530	96	95 - 105	P	01/17/2025	11:21	LB134334
	Zinc	999	1000	100	95 - 105	P	01/17/2025	11:21	LB134334

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Parkway Generation Operating LLC **SDG No.:** Q1097
Contract: PSEG04 **Lab Code:** CHEM **Case No.:** Q1097 **SAS No.:** Q1097
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	22.0	20.0	110	80 - 120	P	01/17/2025	11:26	LB134334
	Nickel	39.2	40.0	98	80 - 120	P	01/17/2025	11:26	LB134334
	Zinc	42.2	40.0	106	80 - 120	P	01/17/2025	11:26	LB134334

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Parkway Generation Operating LLC **SDG No.:** Q1097

Contract: PSEG04 **Lab Code:** CHEM **Case No.:** Q1097 **SAS No.:** Q1097

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1250	1250	100	90 - 110	P	01/17/2025	11:56	LB134334
	Nickel	2470	2500	99	90 - 110	P	01/17/2025	11:56	LB134334
	Zinc	2560	2500	102	90 - 110	P	01/17/2025	11:56	LB134334
CCV02	Copper	1250	1250	100	90 - 110	P	01/17/2025	12:47	LB134334
	Nickel	2440	2500	98	90 - 110	P	01/17/2025	12:47	LB134334
	Zinc	2520	2500	101	90 - 110	P	01/17/2025	12:47	LB134334
CCV03	Copper	1210	1250	96	90 - 110	P	01/17/2025	13:36	LB134334
	Nickel	2360	2500	94	90 - 110	P	01/17/2025	13:36	LB134334
	Zinc	2440	2500	98	90 - 110	P	01/17/2025	13:36	LB134334
CCV04	Copper	1240	1250	100	90 - 110	P	01/17/2025	14:27	LB134334
	Nickel	2440	2500	97	90 - 110	P	01/17/2025	14:27	LB134334
	Zinc	2520	2500	101	90 - 110	P	01/17/2025	14:27	LB134334
CCV05	Copper	1220	1250	98	90 - 110	P	01/17/2025	15:34	LB134334
	Nickel	2390	2500	96	90 - 110	P	01/17/2025	15:34	LB134334
	Zinc	2430	2500	97	90 - 110	P	01/17/2025	15:34	LB134334
CCV06	Copper	1230	1250	98	90 - 110	P	01/17/2025	16:16	LB134334
	Nickel	2410	2500	96	90 - 110	P	01/17/2025	16:16	LB134334
	Zinc	2460	2500	98	90 - 110	P	01/17/2025	16:16	LB134334
CCV07	Copper	1220	1250	98	90 - 110	P	01/17/2025	17:11	LB134334
	Nickel	2400	2500	96	90 - 110	P	01/17/2025	17:11	LB134334
	Zinc	2450	2500	98	90 - 110	P	01/17/2025	17:11	LB134334
CCV08	Copper	1240	1250	99	90 - 110	P	01/17/2025	17:52	LB134334
	Nickel	2420	2500	97	90 - 110	P	01/17/2025	17:52	LB134334
	Zinc	2490	2500	99	90 - 110	P	01/17/2025	17:52	LB134334