

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Parkway Generation Operating LLC **SDG No.:** Q1097
Contract: PSEG04 **Lab Code:** CHEM **Case No.:** Q1097 **SAS No.:** Q1097
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	20.6	2.0	1030	-18	22	01/17/2025	11:39	LB134334
	Nickel	2.52	2.0	126	-38	42	01/17/2025	11:39	LB134334
	Zinc	3.10			-40	40	01/17/2025	11:39	LB134334
ICSAB01	Copper	496	511	97	434	588	01/17/2025	11:43	LB134334
	Nickel	974	954	102	810	1100	01/17/2025	11:43	LB134334
	Zinc	1050	952	110	809	1095	01/17/2025	11:43	LB134334
ICSA	Copper	26.1	2.0	1305	-18	22	01/17/2025	11:47	LB134334
	Nickel	5.63	2.0	282	-38	42	01/17/2025	11:47	LB134334
	Zinc	-6.14			-40	40	01/17/2025	11:47	LB134334
ICSAB	Copper	534	511	104	434	588	01/17/2025	11:52	LB134334
	Nickel	993	954	104	810	1100	01/17/2025	11:52	LB134334
	Zinc	1030	952	108	809	1095	01/17/2025	11:52	LB134334