

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Parkway Generation Operating LLC</u>	level: <u>low</u>	sdg no.: <u>Q1097</u>
contract: <u>PSEG04</u>	lab code: <u>CHEM</u>	case no.: <u>Q1097</u> sas no.: <u>Q1097</u>
matrix: <u>Water</u>	sample id: <u>Q1097-01</u>	client id: <u>402MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q1097-01MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	146		10.0	U	150	97		P
Nickel	ug/L	75 - 125	249		9.50	J	250	96		P
Zinc	ug/L	75 - 125	120		17.1	J	100	103		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Parkway Generation Operating LLC</u>	level: <u>low</u>	sdg no.: <u>Q1097</u>
contract: <u>PSEG04</u>	lab code: <u>CHEM</u>	case no.: <u>Q1097</u> sas no.: <u>Q1097</u>
matrix: <u>Water</u>	sample id: <u>Q1097-01</u>	client id: <u>402MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q1097-01MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	144		10.0	U	150	96		P
Nickel	ug/L	75 - 125	246		9.50	J	250	95		P
Zinc	ug/L	75 - 125	121		17.1	J	100	104		P