

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1100

Project: Monthly 2025

Run No.: LB134301

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134301BS							
BOD5	mg/L	198	193		97	1	84.6-115.4	01/16/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1100

Project: Monthly 2025

Run No.: LB134303

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134303BS							
TPH	mg/L	20.0	16.7		84	1	78-114	01/16/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1100

Project: Monthly 2025

Run No.: LB134303

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134303BSD							
TPH	mg/L	20.0	17.7		88	1	78-114	01/16/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1100

Project: Monthly 2025

Run No.: LB134304

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134304BS							
TSS	mg/L	550	510		93	1	90-110	01/16/2025