

Prep Standard - Chemical Standard Summary

Order ID : Q1101

Test : Ammonia,BOD5,Oil and Grease,Phosphorus-Ortho,Phosphorus-Total,TSS

Prepbatch ID : PB166064,PB166065,

Sequence ID/Qc Batch ID: LB134295,LB134299,LB134301,LB134304,LB134305,LB134307,

Standard ID :

EP2577,WP100827,WP100828,WP108840,WP109922,WP110019,WP110149,WP110150,WP110335,WP110380,WP110386,WP110400,WP110401,WP110416,WP110587,WP110588,WP110826,WP111317,WP111318,WP111323,WP111325,WP111385,WP111404,WP111405,WP111406,WP111407,WP111408,WP111410,WP111411,WP111412,WP111413,WP111414,WP111415,WP111419,WP111420,WP111423,WP111424,WP111425,WP111426,WP111427,WP111428,WP111429,WP99896,

Chemical ID :

E3551,M5673,M6069,M6121,W1992,W1993,W2306,W2606,W2650,W2653,W2654,W2664,W2666,W2699,W2700,W2708,W2783,W2788,W2845,W2858,W2898,W2979,W3035,W3059,W3074,W3103,W3105,W3109,W3110,W3112,W3113,W3132,W3133,W3140,W3143,W3144,W3149,W3155,



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2577	01/06/2025	07/01/2025	Rajesh Parikh	Extraction_SC ALE_2 (EX-SC-2)	None	RUPESHKUMAR SHAH 01/06/2025
<u>FROM</u> 4000.00000gram of E3551 = Final Quantity: 4000.000 gram								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
114	hexavalent chromium color reagent	WP100827	02/02/2023	02/09/2023	Rubina Mughal	WETCHEM_S CALE_5 (WC SC-5)	None	Iwona Zarych 02/02/2023
<u>FROM</u> 0.25000gram of W2979 + 50.00000ml of W2783 = Final Quantity: 50.000 ml								

Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3456	Cyanide Intermediate Working Std, 5PPM	WP100828	02/02/2023	02/03/2023	Iwona Zarych	None	WETCHEM_FIPETTE_3 (WC)	Sohil Jodhani 02/07/2023

FROM 0.25000ml of W2898 + 49.75000ml of WP99896 = Final Quantity: 50.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
635	EDTA BUFFER FOR AMMONIA	WP108840	07/26/2024	01/26/2025	Rubina Mughal	WETCHEM_SCALE_5 (WC SC-5)	None	Iwona Zarych 07/26/2024

FROM 5.50000gram of W3113 + 50.00000gram of W3132 + 950.00000ml of W3112 = Final Quantity: 1000.000 ml

Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1211	11 N sulfuric acid	WP109922	09/26/2024	03/26/2025	Iwona Zarych	None	None	Jignesh Parikh
								10/07/2024

FROM 306.00000ml of M5673 + 694.00000ml of W3112 = Final Quantity: 1000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
289	Sodium Hypochlorite for Ammonia	WP110019	10/02/2024	01/31/2025	Rubina Mughal	None	None	Iwona Zarych
								10/04/2024

FROM 50.00000ml of W3112 + 50.00000ml of W3143 = Final Quantity: 100.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
153	Ammonia Stock Std. (1000 ppm)	WP110149	10/11/2024	04/08/2025	Rubina Mughal	WETCHEM_S CALE_5 (WC SC-5)	None	Iwona Zarych 10/14/2024
<u>FROM</u>	3.81900gram of W1993 + 996.18100ml of W3112 = Final Quantity: 1000.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1895	Ammonia Stock Std, 1000PPM-SS	WP110150	10/11/2024	04/08/2025	Rubina Mughal	WETCHEM_S CALE_5 (WC SC-5)	None	Iwona Zarych 10/14/2024
<u>FROM</u> 3.81900gram of W1992 + 996.18100ml of W3112 = Final Quantity: 1000.000 ml								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1597	0.04 N H2SO4	WP110335	10/22/2024	04/22/2025	Rubina Mughal	None	WETCHEM_PIPETTE_3 (WC)	Iwona Zarych 10/22/2024
<u>FROM</u> 1.00000ml of M5673 + 999.00000ml of W3112 = Final Quantity: 1000.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
126	5N sulfuric acid	WP110380	10/24/2024	04/24/2025	Rubina Mughal	None	None	Iwona Zarych 10/24/2024
<u>FROM</u> 140.00000ml of M5673 + 860.00000ml of W3112 = Final Quantity: 1.000 L								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1841	Sulfuric Acid, 1N	WP110386	10/24/2024	04/24/2025	Rubina Mughal	None	WETCHEM_PIPETTE_3 (WC)	Iwona Zarych 10/24/2024
<u>FROM</u>	2.80000ml of M5673 + 97.20000ml of W3112 = Final Quantity: 100.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
115	Phosphate Stock Std. (50 ppm)	WP110400	10/24/2024	04/23/2025	Rubina Mughal	WETCHEM_S CALE_5 (WC SC-5)	None	Iwona Zarych 10/25/2024
<u>FROM</u> 0.11000gram of W2699 + 500.00000ml of W3112 = Final Quantity: 500.000 ml								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2790	Phosphate Stock std, 50PPM-SS	WP110401	10/24/2024	04/24/2025	Rubina Mughal	WETCHEM_S CALE_5 (WC SC-5)	None	Iwona Zarych 10/25/2024
<u>FROM</u>	0.11000gram of W2708 + 500.00000ml of W3112 = Final Quantity: 500.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
740	sodium nitroferricyanide for ammonia	WP110416	10/25/2024	04/25/2025	Rubina Mughal	WETCHEM_SCALE_5 (WCS-5)	None	Iwona Zarych 10/25/2024
<u>FROM</u> 0.05000gram of W2666 + 99.95000ml of W3112 = Final Quantity: 100.000 ml								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
648	Ammonium molybdate solution	WP110587	11/07/2024	05/07/2025	Niha Farheen Shaik	WETCHEM_S CALE_5 (WC SC-5)	None	Jignesh Parikh 11/07/2024
<u>FROM</u>	20.00000gram of W2664 + 480.00000ml of W3112 = Final Quantity: 500.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
588	Potassium Antimonyl Tartrate	WP110588	11/07/2024	05/07/2025	Niha Farheen Shaik	WETCHEM_S CALE_5 (WC SC-5)	None	Jignesh Parikh 11/07/2024
<u>FROM</u>	1.37150gram of W2306 + 500.00000ml of W3112 = Final Quantity: 500.000 ml							

Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
229	1:1 HCL	WP110826	11/22/2024	05/13/2025	Jignesh Parikh	None	None	Iwona Zarych
								11/22/2024

FROM 500.00000ml of M6121 + 500.00000ml of W3112 = Final Quantity: 1.000 L

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1796	NaOH, 0.1N	WP111317	01/09/2025	07/09/2025	Rubina Mughal	WETCHEM_S CALE_7 (WC SC-6)	None	Iwona Zarych
								01/09/2025

FROM 4.00000gram of W3113 + 996.00000ml of W3112 = Final Quantity: 1000.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1471	NaOH Solution, 6N	WP111318	01/09/2025	07/09/2025	Rubina Mughal	WETCHEM_SCALE_7 (WCS-6)	None	Iwona Zarych 01/09/2025
<u>FROM</u> 240.00000gram of W3113 + 760.00000ml of W3112 = Final Quantity: 1000.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1571	Sodium hydroxide, 1N	WP111323	01/09/2025	07/09/2025	Rubina Mughal	WETCHEM_SCALE_8 (WC SC-7)	None	Iwona Zarych 01/09/2025
<u>FROM</u> 4.00000gram of W3113 + 96.00000ml of W3112 = Final Quantity: 100.000 ml								

Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1494	BORATE BUFFER	WP111325	01/09/2025	01/10/2025	Rubina Mughal	None	None	Iwona Zarych
								01/09/2025

FROM 100.00000L of W3112 + 9.50000gram of W2700 + 88.00000ml of WP111317 = Final Quantity: 100.000 L

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
290	Phenol reagent for Ammonia	WP111385	01/13/2025	07/13/2025	Rubina Mughal	WETCHEM_SCALE_8 (WC SC-7)	None	Iwona Zarych
								01/13/2025

FROM 3.20000gram of W3113 + 8.30000gram of W2858 + 88.80000ml of W3112 = Final Quantity: 100.000 ml

Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
122	calibration std. 0 ppm	WP111404	01/15/2025	01/22/2025	Niha Farheen Shaik	None	None	Iwona Zarych
								01/16/2025

FROM 50.00000ml of W3112 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
121	calibration std. phosphate 0.05 ppm	WP111405	01/15/2025	01/22/2025	Niha Farheen Shaik	None	WETCHEM_FIPETTE_3	Iwona Zarych
							(WC)	01/16/2025

FROM 99.90000ml of W3112 + 0.10000ml of WP110400 = Final Quantity: 100.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
120	calibration std. phosphate 0.1 ppm	WP111406	01/15/2025	01/22/2025	Niha Farheen Shaik	None	WETCHEM_PIPETTE_3	Iwona Zarych
<p>(WC)</p> <p>FROM 99.80000ml of W3112 + 0.20000ml of WP110400 = Final Quantity: 100.000 ml</p>								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
119	calibration std. phosphate 0.3 ppm	WP111407	01/15/2025	01/22/2025	Niha Farheen Shaik	None	WETCHEM_PIPETTE_3 (WC)	Iwona Zarych 01/16/2025
<u>FROM</u> 99.40000ml of W3112 + 0.60000ml of WP110400 = Final Quantity: 100.000 ml								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
118	calibration std. phosphate 0.5 ppm	WP111408	01/15/2025	01/22/2025	Niha Farheen Shaik	None	WETCHEM_FIPETTE_3 (WC)	Iwona Zarych 01/16/2025
<u>FROM</u> 99.00000ml of W3112 + 1.00000ml of WP110400 = Final Quantity: 100.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
117	calibration std. phosphate 1 ppm	WP111410	01/15/2025	01/22/2025	Niha Farheen Shaik	None	WETCHEM_PIPETTE_3 (WC)	Iwona Zarych 01/16/2025
<u>FROM</u> 98.00000ml of W3112 + 2.00000ml of WP110400 = Final Quantity: 100.000 ml								

Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
124	phosphate CCV std.	WP111411	01/15/2025	01/22/2025	Niha Farheen Shaik	None	WETCHEM_F IPETTE_3 (WC)	Iwona Zarych 01/16/2025
FROM 99.00000ml of W3112 + 1.00000ml of WP110400 = Final Quantity: 100.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3805	Phosphate ICV-LCS Std	WP111412	01/15/2025	01/22/2025	Niha Farheen Shaik	None	WETCHEM_F IPETTE_3 (WC)	Iwona Zarych 01/16/2025
FROM 99.00000ml of W3112 + 1.00000ml of WP110401 = Final Quantity: 100.000 ml								

Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
590	Ascorbic Acid	WP111413	01/15/2025	01/22/2025	Niha Farheen Shaik	WETCHEM_SCALE_5 (WC SC-5)	None	Iwona Zarych
								01/16/2025

FROM 0.52800gram of W3074 + 30.00000ml of W3112 = Final Quantity: 30.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
658	Combined reagent	WP111414	01/15/2025	01/16/2025	Niha Farheen Shaik	None	Glass Pipette-A	Iwona Zarych
								01/16/2025

FROM 15.00000ml of WP110587 + 30.00000ml of WP111413 + 5.00000ml of WP110588 + 50.00000ml of WP110380 = Final Quantity: 100.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1213	Phenolphthalein indicator	WP111415	01/15/2025	06/04/2025	Niha Farheen Shaik	WETCHEM_SCALE_5 (WC-5)	None	Iwona Zarych 01/16/2025
<u>FROM</u>	0.10000gram of W2650 + 50.00000ml of W2788 + 50.00000ml of W3112 = Final Quantity: 100.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1322	Ammonia Intermediate Std, 50PPM	WP111419	01/16/2025	02/16/2025	Rubina Mughal	None	WETCHEM_PIPETTE_3 (WC)	Iwona Zarych 01/16/2025
<u>FROM</u> 95.00000ml of W3112 + 5.00000ml of WP110149 = Final Quantity: 100.000 ml								

Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1639	Ammonia Intermediate Std-Second source, 50PPM	WP111420	01/16/2025	02/16/2025	Rubina Mughal	None	WETCHEM_F IPETTE_3 (WC)	Iwona Zarych 01/16/2025
FROM 95.00000ml of W3112 + 5.00000ml of WP110150 = Final Quantity: 100.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
275	Ammonia Calibration Std. (2 ppm)	WP111423	01/16/2025	01/17/2025	Rubina Mughal	None	WETCHEM_F IPETTE_3 (WC)	Iwona Zarych 01/16/2025
FROM 48.00000ml of W3112 + 2.00000ml of WP111419 = Final Quantity: 50.000 ml								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
285	Ammonia CCV Std. (1 ppm)	WP111424	01/16/2025	01/17/2025	Rubina Mughal	None	WETCHEM_PIPETTE_3	Iwona Zarych
<p>(WC)</p> <p>FROM 49.00000ml of W3112 + 1.00000ml of WP111419 = Final Quantity: 50.000 ml</p>								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
286	Ammonia ICV Std. (1 ppm)	WP111425	01/16/2025	01/17/2025	Rubina Mughal	None	WETCHEM_PIPETTE_3 (WC)	Iwona Zarych 01/16/2025
<u>FROM</u> 49.00000ml of W3112 + 1.00000ml of WP111420 = Final Quantity: 50.000 ml								

Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
127	BOD Dilution fluid	WP111426	01/16/2025	01/17/2025	Rubina Mughal	None	None	Iwona Zarych
								01/16/2025

FROM 18.00000L of W3112 + 3.00000PILLOW of W3144 = Final Quantity: 18.000 L

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
129	Glutamic acid-glucose mix for BOD	WP111427	01/16/2025	01/17/2025	Rubina Mughal	WETCHEM_SCALE_7 (WC SC-6)	None	Iwona Zarych
								01/16/2025

FROM 0.15000gram of W2653 + 0.15000gram of W2654 + 1000.00000ml of W3112 = Final Quantity: 1000.000 ml

Wet Chemistry STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
128	polyseed seed control	WP111428	01/16/2025	01/17/2025	Rubina Mughal	None	None	Iwona Zarych
								01/16/2025

FROM 1.00000PILLOW of W3059 + 300.00000ml of WP111426 = Final Quantity: 300.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
658	Combined reagent	WP111429	01/16/2025	01/17/2025	Niha Farheen Shaik	None	Glass Pipette-A	Iwona Zarych
								01/16/2025

FROM 15.00000ml of WP110587 + 30.00000ml of WP111413 + 5.00000ml of WP110588 + 50.00000ml of WP110380 = Final Quantity: 100.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
11	Sodium hydroxide absorbing solution 0.25 N	WP99896	11/15/2022	05/15/2023	Jignesh Parikh	WETCHEM_SCALE_4 (WCS-4)	None	Iwona Zarych 11/15/2022
FROM 21.00000L of W2606 + 210.00000gram of W2845 = Final Quantity: 21.000 L								

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	07/01/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	23D2462010	03/20/2028	09/21/2023 / mohan	09/05/2023 / mohan	M5673

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	140440 / TEST PAPERS,PH,0-2.5,.2SENSI, 100PK	80A0441	02/29/2028	09/03/2024 / jignesh	08/19/2024 / Jaswal	M6069

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9530-33 / Hydrochloric Acid, Instra-Analyzed (cs/6x2.5L)	0000275677	05/13/2025	11/13/2024 / Eman	10/13/2024 / Eman	M6121

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J0660-1 / AMMONIUM CHLORIDE, ACS, 500G	WL13B	04/08/2025	04/08/2015 / apatel	04/08/2015 / apatel	W1992

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J0660-1 / AMMONIUM CHLORIDE, ACS, 500G	XE09B	04/08/2025	04/08/2015 / apatel	04/08/2015 / apatel	W1993

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	A1561-500GM / POTASSIUM ANTIMONY TARTRATE TRIHYDRATE, 500G	2GH0057	12/11/2027	12/11/2017 / apatel	12/11/2017 / apatel	W2306

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	10/24/2024	10/24/2019 / apatel	10/24/2019 / apatel	W2606

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J2870-1 / PHENOLPHTHALEIN, INDICATOR F/TITRATION, 500G	0000235350	06/04/2025	01/31/2020 / AMANDEEP	01/20/2020 / apatel	W2650

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AC156212500 / GLUTAMIC ACID BIOCHEM REG, 250G	A0405990	01/24/2030	01/24/2020 / apatel	01/24/2020 / apatel	W2653

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	D16-500 / DEXTROSE ANHYDROUS ACS REAGENT, 500G(New)	186122A	01/24/2030	01/24/2020 / apatel	01/24/2020 / apatel	W2654

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J07716-1 / Ammonium Molybdate 500G	0000234410	02/11/2026	02/10/2020 / AMANDEEP	01/31/2020 / apatel	W2664

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	87683 / Sodium Nitroferricyanide 250g	W12F013	02/10/2030	02/10/2020 / apatel	02/10/2020 / apatel	W2666

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J3246-1 / POTAS PHOSPHATE, MONO, CRYST, ACS, 500G	04/2019-20	04/23/2025	04/23/2020 / apatel	03/11/2020 / apatel	W2699

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J3568-1 / Sodium Borate, 500 gms	2019111354	04/23/2025	04/23/2020 / apatel	03/11/2020 / apatel	W2700

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J3246-1 / POTAS PHOSPHATE, MONO, CRYST, ACS, 500G	99/2019-20	05/05/2025	05/05/2020 / apatel	05/05/2020 / apatel	W2708

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	0000263246	06/17/2023	12/23/2020 / ketankumar	12/23/2020 / ketankumar	W2783

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC16721-3 / Isopropanol, 99%	C20F23007	06/23/2025	12/30/2020 / apatel	12/30/2020 / apatel	W2788

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19510-7 / Sodium Hydroxide Pellets 12 Kg	21C2456604	01/31/2024	03/30/2022 / JIGNESH	06/24/2021 / apatel	W2845

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	P1060-10 / PHENOL, ACS, 500G	M13H048	01/07/2026	07/07/2021 / apatel	07/07/2021 / apatel	W2858

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Supelco	90157 / Cyanide Standard, 1000ppm from Supelco	HC03107133	06/30/2023	01/24/2022 / apatel	01/24/2022 / apatel	W2898

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	31390 / 1,5-Diphenylcarbazine	MKCR6636	12/09/2027	12/09/2022 / lwona	12/09/2022 / lwona	W2979

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	BDH0214-500G / Ammonium Persulfate Crystal, 500g	MKCR9319	06/30/2028	03/05/2024 / lwona	06/06/2023 / lwona	W3035

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	136742-80 / POLYSEED	152305	05/30/2025	02/15/2024 / Rubina	10/18/2023 / lwona	W3059

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J0938-7 / Ascorbic Acid, 500 gms	MKCS4627	09/30/2025	01/16/2024 / lwona	01/16/2024 / lwona	W3074

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	4620-32 / MANGANOUS SULFATE SOLUTION-364	2403J02	03/31/2026	04/22/2024 / lwona	04/22/2024 / lwona	W3103

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL69870-8 / SODIUM THIOSULFATE,0.025N,4LIT RE	4403S13	09/30/2025	04/22/2024 / lwona	04/22/2024 / lwona	W3105

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL04100-4 / Alkaline Iodide Azide, 1 L	1405D67	04/30/2026	05/23/2024 / lwona	05/23/2024 / lwona	W3109

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	235898	02/28/2029	06/27/2024 / jignesh	06/26/2024 / jignesh	W3110

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19510-7 / Sodium Hydroxide Pellets 12 Kg	23B1556310	12/31/2025	07/08/2024 / lwona	07/08/2024 / lwona	W3113

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC05050-1 / EDTA, disodium salt, dihydrate 1 lb	2ND0156	07/10/2026	07/26/2024 / lwona	07/26/2024 / lwona	W3132

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	140476 / Test Paper,PH Short Range 9.0/10.0	L23	08/22/2029	08/22/2024 / lwona	08/22/2024 / lwona	W3133

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	140444 / TEST PAPERS,PH 0-14,.5 SENSI,100PK	10D0142	09/17/2029	09/17/2024 / lwona	09/17/2024 / lwona	W3140

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J9416-1 / Sodium Hypochlorite 500 ml	2407F34	01/31/2025	09/30/2024 / lwona	09/30/2024 / lwona	W3143

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
HACH	1486266 / BOD Nutrient Buffer Pillows, 6 mL concentrate to make 6 L, 50/pk	A4169	06/30/2029	11/20/2024 / rubina	10/01/2024 / lwona	W3144

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL70850-8 / Starch Solution, 4L	4408P62	08/31/2026	10/16/2024 / Iwona	10/16/2024 / Iwona	W3149

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	140730 / TEST PAPER,POT.IOD-STRCH,P K100,CS12	14-860	12/02/2029	12/02/2024 / Iwona	12/02/2024 / Iwona	W3155



CERTIFICATE OF ANALYSIS

Printed: 12/8/2017

Page 1 of 1

Customer No : 30017
Order Number : 3008126
Catalog : A1561

Customer : PCI SCIENTIFIC
Delivery # : 58495347
Potassium Antimony Tartrate Trihydrate,
Reagent, ACS

Customer PO : 6035343
Lot : 2GH0057

Chemical Formula : $C_8H_4K_2O_{12}Sb_2 \cdot 3H_2O$
CAS# : 28300-74-5

Formula Weight : 667.87

W2306
received
12/11/17
AB

Test

Limit
Min. Max.

Results

ASSAY ($C_8H_4K_2O_{12}Sb_2 \cdot 3HO$)	99.0 - 103.0 %	101.0 %
TITRATABLE ACID OR BASE	-- 0.020 meq/g	<0.020 meq/g
LOSS ON DRYING	-- 2.7 %	<2.7 %
ARSENIC (As)	-- 0.015 %	<0.015 %
APPEARANCE		WHITE POWDER
DATE OF MANUFACTURE		29-DEC-2015

All pharmaceutical ingredients are tested using current edition of applicable pharmacopeia.

Read and understand label and MSDS/SDS before handling any chemical. All Spectrum's chemicals are for manufacturing, processing, repacking or research purposes by experienced personnel only. The customer must ensure to provide its users adequate hazardous material training and appropriate protective gears before handling our chemicals.

Certificate of Analysis Results Certified By:

Certificate of Analysis



Date of Release: 12/18/2013

Product: Ammonium Chloride GR ACS

Catalog No.: AX1270 all
size codes

Grade: Meets ACS Specifications

CAS #: 12125-02-9

Country of Origin: India

FW: 53.49

Lot No.: WL13B



Characteristic	Requirement		Results	UOM
	Minimum	Maximum		
Assay (argentometric)	99.5		99.9	%
Calcium (Ca)		0.001	0.0001	%
Form	White crystals		White crystals	
Heavy metals (as Pb)		5	5	ppm
Identification	To pass test		Passes	
Insoluble matter		0.005	0.002	%
Iron (Fe)		2	2	ppm
Loss on drying (105 C)		0.5	0.21	%
Magnesium (Mg)		5	0.6	ppm
pH of a 5% solution at 25 C	4.5	5.5	4.76	
Phosphate (PO4)		2	2	ppm
Residue after ignition		0.01	0.002	%
Sulfate (SO4)		0.002	0.002	%

Joe Schoellkopff

Quality Control Manager

This document has been produced electronically and is valid without a signature.

Certificate of Analysis



Date of Release: 5/12/2014

Product: Ammonium Chloride GR ACS

Catalog No.: AX1270 all
size codes

Grade: Meets ACS Specifications

CAS #: 12125-02-9

Country of Origin: India

FW: 53.49

Lot No.: XE09B



Characteristic	Requirement		Results	UOM
	Minimum	Maximum		
Assay (argentometric)	99.5		99.8	%
Calcium (Ca)		0.001	0.0001	%
Form	White crystals		White crystals	
Heavy metals (as Pb)		5	5	ppm
Identification	To pass test		Passes	
Insoluble matter		0.005	0.002	%
Iron (Fe)		2	2	ppm
Loss on drying (105 C)		0.5	0.22	%
Magnesium (Mg)		5	0.7	ppm
pH of a 5% solution at 25 C	4.5	5.5	4.95	
Phosphate (PO4)		2	2	ppm
Residue after ignition		0.01	0.002	%
Sulfate (SO4)		0.002	0.002	%

Joe Schoellkopf

Quality Control Manager

This document has been produced electronically and is valid without a signature.

Subject to Vadodara Jurisdiction



CHAMPA PURIE-CHEM INDUSTRIES

ISO 9001 : 2015 CERTIFIED COMPANY

Importers Exporters Manufacturers & Marketing of Fine Chemicals & Pharmaceuticals

262-263, G.I.D.C. Estate,
Makarpura,
Vadodara - 390 010.
Gujarat - INDIA.

Phone : (F) +91-265-2633314 / 2643723
Fax : (F) +91-265-2638038
E-mail : info@cpclndia.com
Web : www.cpcindia.com

CERTIFICATE OF ANALYSIS

PRODUCT : POTASSIUM PHOSPHATE MONOBASIC Anhy. - ACS		
CERTIFICATE NO	: 04/2019-20	DATE 13-05-2019
Date of receipt of sample	: 29.04.2019	Quantity : 1000 KGS.
Batch No. /Lot No.	: 04/2019-20	
Mfg. Date	: April-2019	
1. Characteristic : A White powder		
2. Identification : Positive		
	RESULT OBTAINED	LIMITS
3. Clarity and colour of solution : 10% solution is clear and colourless		
4. Assay (on dry basis)	99.35%	Min.99.00%
5. PH (5% solution)	4.28	4.1-4.5
6. Loss on Drying	0.06%	Max 0.2%
7. Heavy Metals	0.0004%	Max.0.001%
8. iron	0.001%	Max 0.002%
9. Sulphate	0.0015%	Max. 0.003%
10. Chloride	0.0005%	Max.0.001%
11. Insoluble Matter	0.002%	Max. 0.01%
12. Sodium	0.0038%	Max. 0.005%
The sample does comply with specification as per Above.		
Analysed by <u>J. A. PATHAK</u>		Quality Control Department



Certificate of Analysis

1.19533.0500 Cyanide standard solution traceable to SRM from NIST $K_2[Zn(CN)_4]$ in H_2O
1000 mg/l CN Certipur®
Batch HC03107133

Batch Values

Concentration	β (CN ⁻)	1002	mg/l
---------------	----------------------------	------	------

Determination method: Argentometric titration.

The content of this solution was determined with silver nitrate standard solution (article number 1.09081) standardized against volumetric standard sodium chloride (article number 1.02406). The expanded measurement uncertainty is $\pm 0.7\%$ ($k=2$ coverage factor for 95% coverage probability). The certified value is traceable to primary standard NIST SRM 999c (NIST: National Institute of Standards and Technology, USA) by means of volumetric standard sodium chloride, measured in the accredited calibration laboratory of Merck KGaA, Darmstadt, Germany in accordance to DIN EN ISO/IEC 17025.

Date of release (DD.MM.YYYY) 02.07.2020

Minimum shelf life (DD.MM.YYYY) 30.06.2023

Ayfer Yildirim

Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.

Ammonium Molybdate, 4-Hydrate, Crystal
BAKER ANALYZED® A.C.S. Reagent

(ammonium heptamolybdate, tetrahydrate)



Material No.: 0716-01
Batch No.: 0000234410
Manufactured Date: 2019/02/13
Retest Date: 2026/02/11
Revision No: 1

Certificate of Analysis

Meets ACS Reagent Chemical Requirements,

Test	Specification	Result
Assay (as MoO ₃)	81.0 – 83.0 %	81.4
ACS – Insoluble Matter	<= 0.005 %	< 0.001
Chloride (Cl)	<= 0.002 %	< 0.002
Nitrate (NO ₃)	Passes Test	PT
Arsenate, Phosphate and Silicate (as SiO ₂)	<= 0.001 %	< 0.001
ACS – Phosphate (PO ₄)	<= 5 ppm	< 5
Sulfate (SO ₄)	<= 0.02 %	< 0.02
Heavy Metals (as Pb)	<= 0.001 %	< 0.001
Magnesium (Mg)	<= 0.005 %	< 0.001
Potassium (K)	<= 0.01 %	< 0.01
Sodium (Na)	<= 0.01 %	<0.001

For Laboratory, Research or Manufacturing Use
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: US
Packaging Site: Paris Mfg Ctr & DC

James Ethier
Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Phenolphthalein, Powder
BAKER ANALYZED® A.C.S. Reagent



Material No.: 2870-01
Batch No.: 0000235350
Manufactured Date: 2018/06/06
Retest Date: 2025/06/04
Revision No: 1

Certificate of Analysis

Meets ACS Reagent Chemical Requirements,

Test	Specification	Result
ACS – Clarity of Solution	Passes Test	PT
Visual Transition Interval – pH...8.0 (Colorless)	Passes Test	PT
Visual Transition Interval – pH...10.0 (Red)	Passes Test	PT

For Laboratory, Research or Manufacturing Use

Country of Origin: CN
Packaging Site: Paris Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Acetone
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9254-03
Batch No.: 0000263246
Manufactured Date: 2020/06/17
Expiration Date: 2023/06/17
Revision No: 1

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.7
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0000 ppm	0.1000
Substances Reducing Permanganate	Passes Test	PT
Titration Acid (μeq/g)	<= 0.3	0.1
Titration Base (μeq/g)	<= 0.6	< 0.1
Water (H ₂ O)	<= 0.5 %	0.3
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	< 1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	5

For Laboratory, Research or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: US
Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

W2858 Received by AP on 07/07/2021

Product No.: 33213
Product: Phenol, ACS, 99+%, stab.
Lot No.: M13H048

Test	Limits	Results
Assay	99.0 % min	99.8 %
Freezing point	40.5°C min	40.5 °C
Clarity of solution	To pass test	Passes
Residue after evaporation	0.05 % max	< 0.05 %
Water	0.5 % max	0.2 %

Retest date: January 7, 2026

Order our products online [alfa.com](https://www.alfa.com)**This document has been electronically generated and does not require a signature.**

This is to certify that units of the lot number above were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the purchaser, formulator or those performing further manufacturing to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The above information is the actual analytical results obtained.

W2666 Recived on 02/10/2020 by AP

Product No.: 87683

Product: Sodium pentacyanonitrosylferrate(III) dihydrate, ACS, 99.0-102.0%


Lot No.: W12F013

Test	Limits	Results
Assay	99.0 - 102.0 %	99.67 %
Insoluble	0.01 % max	0.0079 %
Chloride	0.02 % max	Not detected
Sulfate	To pass test	Passes test
Aqueous solubility	To pass test	Passes test
Limit on Ferricyanide	To pass test	Passes test
Limit on Ferrocyanide	To pass test	Passes test

Order our products online [alfa.com](https://www.alfa.com)

This document has been electronically generated and does not require a signature.

This is to certify that units of the lot number above were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the purchaser, formulator or those performing further manufacturing to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The above information is the actual analytical results obtained.



ACROS ORGANICS part of Thermo Fisher Scientific

ACROS ORGANICS

Version	0
Molecular weight	147.13
Molecular formula	C ₅ H ₉ N O ₄
CAS No	56-86-0
Linear formula	HO ₂ CCH ₂ CH ₂ CH(NH ₂)CO ₂ H
Flash point (°C)	

Certificate of Analysis

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Acros Organics expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to human or animals. It is the responsibility of the purchaser, formulator or those performing further manufacturing to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	15621	Quality Test / Release Date	13 March 2019
Lot Number	A0405990	Suggested Retest Date	March 2022
Description	L(+)-Glutamic acid, 99%		
Country of Origin	CHINA		
Declaration of Origin	plant		

Origin Comment	The product is made by fermentation of sugar molasses
-----------------------	---

Result Name	Specifications	Test Value
Appearance (Color)	White	White
Appearance (Form)	Powder	Powder
Infrared spectrum	Conforms	Conforms
Titration with NaOH	98.5 to 100.5 % (On dried substance)	99.32 % (On dried substance)
Loss on drying	≤0.5 % (105°C, 3 hrs)	0.002 % (105°C, 3 hrs)
Heavy metals (as Pb)	≤10 ppm	≤10 ppm
Sulfated ash	≤0.1 %	0.08 %
Other amino acids	not detectable	not detectable
Specific optical rotation	+30.5° to +32.5° (20°C, 589 nm) (on dried substance)	+32° (20°C, 589 nm) (on dried substance)
Specific optical rotation	(c=10, 2N HCl)	(c=10, 2N HCl)
Chloride (Cl)	≤200 ppm	≤200 ppm
Iron (Fe)	≤30 ppm	≤10 ppm
Sulfate (SO ₄)	≤300 ppm	≤200 ppm
Ammonium (NH ₄)	≤200 ppm	≤200 ppm
Arsenic oxide (As ₂ O ₃)	≤1 ppm	≤1 ppm



A handwritten signature in black ink, which appears to read "L. Van den Broek".

L. Van den Broek, QA Manager

Issued: 24 January 2020

Acros Organics

ENA23, zone 1, nr 1350, Janssen Pharmaceuticaaan 3a, B-2440 Geel, Belgium

Tel +32 14/57.52.11 - Fax +32 14/59.34.34 Internet: <http://www.acros.com>

1 Reagent Lane, Fair Lawn, NJ 07410, USA Fax 201-796-1329



CERTIFICATE OF ANALYSIS

Product Name ISOPROPYL ALCOHOL, 99%
Grade Meets ACS/USP/NF Monographs
Catalog # 231000099, zp231000099
Lot # C20F23007
Date of Manufacture: 06/23/20 **W2788 Received on 12/30/2020 by AP**
Recommended Retest Date: Five Years from Date of Manufacture

TEST	MONO GRAPH	SPECIFICATION	RESULT
Assay (corrected for water)	USP	99.0% min	99.92%
Assay (corrected for water)	ACS	99.5% min	
Solubility in water	ACS ⁺	To Pass Test	Pass
Appearance	ACS ⁺	Clear, colorless liquid	Pass
Color, APHA	ACS	10 max	1
Limit of Nonvolatile Residue	USP ⁺	NMT 2.5 mg (0.005%)	0.1 mg
Residue after Evaporation	ACS ⁺	0.001% max	< 0.001%
Specific Gravity	USP	0.783 - 0.787 @25°C	0.783
Identification A - Infrared Absorption	USP	To Pass Test	Pass
Identification B	USP	To Pass Test	Pass
Refractive Index @ 20°C	USP	1.376-1.378	1.377
Acidity	USP ⁺	NMT 0.70 ml of 0.020N NaOH is required	0.30 mL
Titration Acid or Base	ACS ⁺	0.0001 meq/g max	0.0001 meq/g
Carbonyl Compounds	ACS	Propionaldehyde 0.002% max	< 0.002%
		Acetone 0.002% max	None Detected
Limit of Volatile Impurities	USP	Diethyl Ether NMT 0.1%	< 0.1%
		Acetone NMT 0.1%	None Detected
		Diisopropyl Ether NMT 0.1%	< 0.1%
		n-Propyl Alcohol NMT 0.1%	< 0.1%
		2-Butanol NMT 0.1%	< 0.1%
		Total NMT 1.0%	< 0.1%
Water, wt%	ACS	NMT 0.2%	0.05%
Water Determination	USP	NMT 0.5%	

⁺This test is performed quarterly

Certification and Compliance Statements

This lot of Isopropyl Alcohol complies with all of the current requirements listed in the United States Pharmacopeia, American Chemical Society monographs and the National Formulary.

No chemicals whatsoever are used as solvents at any point in the manufacture, processing or packaging of Isopropyl Alcohol. Only Class 2 and Class 3 residual solvents may appear as impurities / related substances / low level contaminants in IPA. Concentration of Class 2 Option 1 and Class 3 residual solvents is below limits in the current USP/NF General Chapter <467>.

This product is not derived, nor does it come in contact with, any materials derived from bovine or other animal sources.

This product is for further commercial manufacturing, laboratory or research use, and may be used as an excipient or a process solvent for pharmaceutical purposes. It is not intended for use as an active ingredient in drug manufacturing nor as a medical device or disinfectant. Appropriate/legal use of this product is the responsibility of the user.

Approved by: D. Simoncelli, Quality Control Chemist

Date of Approval: 06/23/2020




W 3035
rec. 6/6/23 12

Product Name:


Certificate of Analysis

Ammonium persulfate - ACS reagent, $\geq 98.0\%$

Product Number: 248614
Batch Number: MKCR9319
Brand: SIGALD
CAS Number: 7727-54-0
MDL Number: MFCD00003390
Formula Weight: 228.20 g/mol
Quality Release Date: 13 OCT 2022



Test	Specification	Result
Appearance (Color)	White to Off White	White
Appearance (Form)	Powder or Crystals or Granules or Chunks	Crystals
ICP Major Analysis	Confirmed	Confirmed
Confirms Sulfur Component		
Titration by KMNO ₄	$\geq 98.0 \%$	100.0 %
Residue on ignition (Ash)	$\leq 0.05 \%$	$< 0.05 \%$
Insoluble Matter	$\leq 0.005 \%$	0.002 %
c = 10 %; In Water		
Chloride and Chlorate (as Cl)	$\leq 0.001 \%$	$< 0.001 \%$
Iron (Fe)	$\leq 0.001 \%$	$< 0.001 \%$
Heavy Metal	$\leq 0.005 \%$	$< 0.001 \%$
as Lead		
Manganese (Mn)	$\leq 0.5 \text{ ppm}$	$< 0.1 \text{ ppm}$
Titrateable Acid (meq/g)	≤ 0.04	< 0.04
Meets ACS Requirements	Current ACS Specification	Conforms


Larry Coers, Director
Quality Control
Milwaukee, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.





CERTIFICATE OF ANALYSIS

PO BOX 130549 Spring, TX 77393
Phone: (281) 298-9410 Fax: (281) 298-9411

FINISHED PRODUCT, LOT NUMBER, MFG. /EXP DATE:

PolySeed® • Part No. P-110 • Lot 152305 • Mfg. Date: 05/2023 • Exp. Date: 05/2025

FORMULATION:

The formulation for this product contains a range of naturally occurring microorganisms, which are known to be non-pathogenic to man or animals.

VIABLE COUNT, FINAL TEST RESULT:

The product has been fully tested in accordance with Finished Product Specifications and contains a minimum viable count of 4.00×10^9 cfu/g.

GLUCOSE/GLUTAMIC-ACID RESULTS:

Tested results within acceptable range 198 ± 30.5 mg/L (167.5 - 228.5 mg/L). GGA Lot# L257-09 – Average Test Result: 203.4

See www.polyseed.com for details.

SEED CONTROL FACTOR:

Tested results within acceptable range 0.6 – 1.0 see www.polyseed.com for details

SALMONELLA TEST RESULT:

The product has been shown to be Salmonella negative using procedures recommended in the Microbiology Laboratory Guidebook, published by the USDA Food Safety and Inspection Service.

The purpose of this document is to assure that the Finished Product conforms to the above specification.

Signature: _____

Quality Control Department

Date: 05/15/2023

POLYSEED.Ref.1.19

Revised Jan 23

InterLab®
International Laboratory Supply



Certificate Of Analysis



Date of Release: 11/14/2019

W2700 Recived by AP on 3/11/2020

Name: **Sodium Borate, Decahydrate**
ACS

Item No: **SX0355 All Sizes**

Lot / Batch No: **2019111354**

Country of Origin: **India**

Item	Specifications	Analysis
Assay (Na ₂ B ₄ O ₇ • 10H ₂ O)	99.5 - 105.0%	101.7%
Calcium (Ca)	0.005% max.	0.003%
Chloride (Cl)	0.001% max.	<0.001%
Color	White	Passes Test
Form	Crystals	Passes Test
Heavy Metals (as Pb)	0.001% max.	<0.001%
Insoluble Matter	0.005% max.	0.002%
Iron (Fe)	5 ppm max.	<5 ppm
pH of a 0.01 M solution at 25C	9.15 - 9.20	9.17
Phosphate (PO ₄)	0.001% max.	<0.001%
Sulfate (SO ₄)	0.005% max.	<0.005%

Joe Schoellkopf

Quality Control Manager

This document has been produced electronically and is valid without a signature .

EMD Millipore is a division of Merck KGaA, Darmstadt, Germany

EMD Millipore Corporation

400 Summit Drive
Burlington, MA 01803
U.S.A.

Form number: 00005624CA, Rev. 2.0



Certificate of Analysis

1 Reagent Lane
Fair Lawn, NJ 07410
201.796.7100 tel
201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120632

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	D16	Quality Test / Release Date	03/19/2019
Lot Number	186122A		
Description	DEXTROSE, ANHYDROUS, A.C.S.		
Country of Origin	United States	Suggested Retest Date	Mar/2022
Chemical Origin	Organic - Plant		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		
Chemical Comment			

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	White, granular powder
TITRATABLE ACID	MEQ/G	<= 0.002	<0.002
STARCH		= PASS TEST	pass test
SPECIFIC ROTATION @ 25 C	DEGREES (+ OR -)	Inclusive Between +52.5 - +53.0	53.0
SULFATE & SULFITE	%	<= 0.005	<0.005
IRON (Fe)	ppm	<= 5	<5
CHLORIDE	%	<= 0.01	<0.01
IGNITION RESIDUE	%	<= 0.02	<0.02
IDENTIFICATION	PASS/FAIL	= PASS TEST	pass test
HEAVY METALS (as Pb)	ppm	<= 5	<5
LOSS ON DRYING @ 105 C	%	<= 0.2	<0.2
INSOLUBLE MATTER	%	<= 0.005	0.002

Jerisa Bailey-Wyche

Quality Assurance Specialist - Certificate of Analysis Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.
If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %
COMMENTS		
QC: PhC Irma Belmares		

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/24/23 E 3551

RC-02-01, Ed. 3

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium

 **avantor**™



Material No.: 9673-33
Batch No.: 23D2462010
Manufactured Date: 2023-03-22
Retest Date: 2028-03-20
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
ACS – Assay (H ₂ SO ₄)	95.0 – 98.0 %	96.1 %
Appearance	Passes Test	Passes Test
ACS – Color (APHA)	≤ 10	5
ACS – Residue after Ignition	≤ 3 ppm	< 1 ppm
ACS – Substances Reducing Permanganate (as SO ₂)	≤ 2 ppm	< 2 ppm
Ammonium (NH ₄)	≤ 1 ppm	1 ppm
Chloride (Cl)	≤ 0.1 ppm	< 0.1 ppm
Nitrate (NO ₃)	≤ 0.2 ppm	< 0.1 ppm
Phosphate (PO ₄)	≤ 0.5 ppm	< 0.1 ppm
Trace Impurities – Aluminum (Al)	≤ 30.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 4.0 ppb	< 2.0 ppb
Trace Impurities – Boron (B)	≤ 10.0 ppb	8.5 ppb
Trace Impurities – Cadmium (Cd)	≤ 2.0 ppb	< 0.3 ppb
Trace Impurities – Chromium (Cr)	≤ 6.0 ppb	< 0.4 ppb
Trace Impurities – Cobalt (Co)	≤ 0.5 ppb	< 0.3 ppb
Trace Impurities – Copper (Cu)	≤ 1.0 ppb	< 0.1 ppb
Trace Impurities – Gold (Au)	≤ 10.0 ppb	0.5 ppb
Heavy Metals (as Pb)	≤ 500.0 ppb	< 100.0 ppb
Trace Impurities – Iron (Fe)	≤ 50.0 ppb	1.3 ppb
Trace Impurities – Lead (Pb)	≤ 0.5 ppb	< 0.5 ppb
Trace Impurities – Magnesium (Mg)	≤ 7.0 ppb	0.8 ppb
Trace Impurities – Manganese (Mn)	≤ 1.0 ppb	< 0.4 ppb
Trace Impurities – Mercury (Hg)	≤ 0.5 ppb	< 0.1 ppb
Trace Impurities – Nickel (Ni)	≤ 2.0 ppb	0.3 ppb
Trace Impurities – Potassium (K)	≤ 500.0 ppb	< 2.0 ppb
Trace Impurities – Selenium (Se)	≤ 50.0 ppb	< 0.1 ppb
Trace Impurities – Silicon (Si)	≤ 100.0 ppb	31.5 ppb
Trace Impurities – Silver (Ag)	≤ 1.0 ppb	< 0.3 ppb

>>> Continued on page 2 >>>

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium



Material No.: 9673-33
Batch No.: 23D2462010

Test	Specification	Result
Trace Impurities – Sodium (Na)	≤ 500.0 ppb	5.4 ppb
Trace Impurities – Strontium (Sr)	≤ 5.0 ppb	< 0.2 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	< 0.8 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	0.4 ppb

For Laboratory, Research, or Manufacturing Use

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality



Certificate of Analysis

Product information

Product	pH-Fix 0.3-2.3
REF	92180
LOT	80A0441
Expiration date:	29.02.2028
Date of examination:	23.01.2024
Gradation:	pH 0.3-0.7-1.0-1.3-1.6-1.9-2.3

Confirmation

Hereby we confirm, that the above mentioned product has successfully passed our quality control system in accordance with ISO 9001 and meets the specific quality criteria.

This document has been produced electronically and is valid without a signature.



Hydrochloric Acid, 36.5-38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis

avantor™



R → 16/13/24
Met dig

M 6121

Material No.: 9530-33
Batch No.: 0000275677
Manufactured Date: 2020/12/16
Retest Date: 2025/12/15
Revision No: 1

Certificate of Analysis

Test	Specification	Result
ACS - Assay (as HCl) (by acid-base titrn)	36.5 - 38.0 %	37.6
ACS - Color (APHA)	<= 10	5
ACS - Residue after Ignition	<= 3 ppm	1
ACS - Specific Gravity at 60°/60°F	1.185 - 1.192	1.190
ACS - Bromide (Br)	<= 0.005 %	< 0.005
ACS - Extractable Organic Substances	<= 5 ppm	1
ACS - Free Chlorine (as Cl ₂)	<= 0.5 ppm	< 0.5
Phosphate (PO ₄)	<= 0.05 ppm	< 0.03
Sulfate (SO ₄)	<= 0.5 ppm	< 0.3
Sulfite (SO ₃)	<= 0.8 ppm	0.3
Ammonium (NH ₄)	<= 3 ppm	< 1
Trace Impurities - Arsenic (As)	<= 0.010 ppm	< 0.003
Trace Impurities - Aluminum (Al)	<= 10.0 ppb	< 0.2
Arsenic and Antimony (as As)	<= 5 ppb	< 3
Trace Impurities - Barium (Ba)	<= 1.0 ppb	< 0.2
Trace Impurities - Beryllium (Be)	<= 1.0 ppb	< 0.2
Trace Impurities - Bismuth (Bi)	<= 10.0 ppb	< 1.0
Trace Impurities - Boron (B)	<= 20.0 ppb	< 5.0
Trace Impurities - Cadmium (Cd)	<= 1.0 ppb	< 0.3
Trace Impurities - Calcium (Ca)	<= 50.0 ppb	29.7
Trace Impurities - Chromium (Cr)	<= 1.0 ppb	< 0.4
Trace Impurities - Cobalt (Co)	<= 1.0 ppb	< 0.3
Trace Impurities - Copper (Cu)	<= 1.0 ppb	< 0.1
Trace Impurities - Gallium (Ga)	<= 1.0 ppb	< 0.2

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Material No.: 9530-33

Batch No.: 0000275677

Test	Specification	Result
Trace Impurities – Germanium (Ge)	≤ 3.0 ppb	< 2.0
Trace Impurities – Gold (Au)	≤ 4.0 ppb	< 0.2
Heavy Metals (as Pb)	≤ 100 ppb	< 50
Trace Impurities – Iron (Fe)	≤ 15.0 ppb	< 1
Trace Impurities – Lead (Pb)	≤ 1.0 ppb	< 0.5
Trace Impurities – Lithium (Li)	≤ 1.0 ppb	0.2
Trace Impurities – Magnesium (Mg)	≤ 10.0 ppb	0.4
Trace Impurities – Manganese (Mn)	≤ 1.0 ppb	< 0.4
Trace Impurities – Mercury (Hg)	≤ 0.5 ppb	0.1
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 5.0
Trace Impurities – Nickel (Ni)	≤ 4.0 ppb	< 0.3
Trace Impurities – Niobium (Nb)	≤ 1.0 ppb	< 0.2
Trace Impurities – Potassium (K)	≤ 9.0 ppb	< 2.0
Trace Impurities – Selenium (Se), For Information Only	ppb	1.0
Trace Impurities – Silicon (Si)	≤ 100.0 ppb	< 10.0
Trace Impurities – Silver (Ag)	≤ 1.0 ppb	< 0.3
Trace Impurities – Sodium (Na)	≤ 100.0 ppb	< 5.0
Trace Impurities – Strontium (Sr)	≤ 1.0 ppb	< 0.2
Trace Impurities – Tantalum (Ta)	≤ 1.0 ppb	< 0.9
Trace Impurities – Thallium (Tl)	≤ 5.0 ppb	< 2.0
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	< 0.8
Trace Impurities – Titanium (Ti)	≤ 1.0 ppb	0.2
Trace Impurities – Vanadium (V)	≤ 1.0 ppb	< 0.2
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	0.3
Trace Impurities – Zirconium (Zr)	≤ 1.0 ppb	< 0.1

For Laboratory, Research or Manufacturing Use


Product Information (not specifications):

Appearance (clear, fuming liquid)

Meets ACS Specifications

Country of Origin: US

Packaging Site: Phillipsburg Mfg Ctr & DC



Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700



CHAMPA PURIE-CHEM INDUSTRIES

ISO 9001 : 2015 CERTIFIED COMPANY

Importers Exporters Manufacturers & Marketing of Fine Chemicals & Pharmaceuticals

262-263, G.I.D.C. Estate,
Makarpura,
Vadodara - 390 010.
Gujarat - INDIA.

Phone : (F) +91-265-2633314 / 2643723
Fax : (F) +91-265-2638036
E-mail : info@cpcindia.com
Web : www.cpcindia.com

W2708 Received on 05/05/20 by AP

CERTIFICATE OF ANALYSIS

PRODUCT	: POTASSIUM PHOSPHATE MONOBASIC Anhy. - ACS	
CERTIFICATE NO	: 99/2019- 20	DATE 26-08-2019
Date of receipt of sample	: 22.08.2019	Quantity : 1000 KGS
Batch No. /Lot No	: 99/2019- 20	
Mfg. Date	: Aug-2019	
1. Characteristic	: A White powder	
2. Identification	: Positive	
	RESULT OBTAINED	LIMITS
3. Clarity and colour of solution	: 10% solution is clear and colourless	
4. Assay (on dry basis)	: 99.27%	Min.99.00%
5. PH (5% solution)	: 4.4	4.1-4.5
6. Loss on Drying	: 0.1%	Max 0.2%
7. Heavy Metals	: 0.0003%	Max.0.001%
8. Iron	: 0.001%	Max 0.002%
9. Sulphate	: 0.001%	Max. 0.003%
10. Chloride	: 0.0005%	Max.0.001%
11. Insoluble Matter	: 0.003%	Max. 0.01%
12. Sodium	: 0.004%	Max. 0.005%
The sample does comply with specification as per Above.		
Analysed by	<u>J.A. PATHAK</u>	Quality Control Department

W 2979

Rec: 12/09/22

exp. 12/09/27

Product Name:

1,5-Diphenylcarbazide - ACS reagent

Product Number:

259225

Batch Number:

MKCR6636

Brand:

SIAL

CAS Number:

140-22-7

MDL Number:

MFCD00003013

Formula:


C₁₃H₁₄N₄O

Formula Weight:

242.28 g/mol


Quality Release Date:

02 JUN 2022



Certificate of Analysis

Test	Specification	Result
Appearance (Color)	Conforms to Requirements	Pink
Off-White to Pink, Light Purple or Tan		
Appearance (Form)	Powder or Chunks	Powder
Melting Point	173.0 - 176.0 °C	173.0 °C
Infrared Spectrum	Conforms to Structure	Conforms
Residue on ignition (Ash)	≤ 0.05 %	0.01 %
15 minutes, 800 Degrees Celsius		
Solubility	Pass	Pass
Sensitivity Test	Pass	Pass
Meets ACS Requirements	Current ACS Specification	Conforms



Larry Coers, Director
Quality Control
Milwaukee, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.




W3074 Rec. on 01/16/24 by IZ

Certificate of Analysis

Product Name:

L-Ascorbic acid - ACS reagent, ≥99%

Product Number: 255564
Batch Number: MKCS4627
Brand: SIAL
CAS Number: 50-81-7
MDL Number: MFCD00064328
Formula: C₆H₈O₆
Formula Weight: 176.12 g/mol
Quality Release Date: 21 NOV 2022
Recommended Retest Date: SEP 2025



Test	Specification	Result
Appearance (Color)	White	White
Appearance (Form)	Conforms to Requirements	Powder
Powder, Crystals, Crystalline Powder, Granules and/or Chunks		
Infrared Spectrum	Conforms to Structure	Conforms
Optical Rotation (+); c = 10%; Water	20.5 - 21.5 deg	20.7 deg
Titration by Iodine	≥ 99.0 %	99.4 %
Residue on Ignition	≤ 0.10 %	0.03 %
Iron (Fe)	≤ 0.001 %	< 0.001 %
Heavy Metals by ICP-OES	≤ 0.002 %	0.001 %
Recommended Retest Period 3 Years	-----	-----
Meets ACS Requirements	Current ACS Specification	Conforms

Larry Coers, Director
Quality Control
Milwaukee, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.





Certificate of Analysis

Manganous Sulfate Solution, 364 g/L**Lot Number:** 2403J02**Product Number:** 4620**Manufacture Date:** MAR 15, 2024**Expiration Date:** MAR 2026

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Manganous Sulfate Monohydrate	10034-96-5	Reagent
Sulfuric Acid	7664-93-9	ACS

Test	Specification	Result
Appearance	Pink liquid	Passed
Assay (by Refractive Index)	360-368 g/L	367 g/L

Specification	Reference
Manganous Sulfate Solution	ASTM (D 888 A)
Manganous Sulfate Solution	ASTM (D 888 A)
Manganous Sulfate Solution	APHA (4500-O E)
Manganous Sulfate Solution	APHA (4500-O F)
Manganous Sulfate Solution	APHA (4500-O D)
Manganous Sulfate Solution	APHA (4500-O E)
Manganous Sulfate Solution	APHA (4500-O F)
Manganous Sulfate Solution	APHA (4500-O D)
Manganous Sulfate Solution	APHA (4500-O C)
Manganous Sulfate Solution	APHA (4500-O C)
Manganous Sulfate Solution	EPA (360.2)
Manganous Sulfate Solution	EPA (360.2)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
4620-32	1 L natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Jose Pena (03/15/2024)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --
Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Certificate of Analysis

Sodium Thiosulfate, 0.0250 Normal (N/40)

Lot Number: 4403S13

Product Number: 7900

Manufacture Date: MAR 29, 2024

Expiration Date: SEP 2025

This product is specially formulated to increase its stability. A preservative is added to prevent bacterial contamination. However, all Sodium Thiosulfate solutions are subject to slow chemical deterioration and should be restandardized periodically.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Thiosulfate Pentahydrate	10102-17-7	ACS
Organic Preservative	Proprietary	
Sodium Carbonate	497-19-8	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Potassium Iodate/Starch)	0.02499-0.02501 N at 20°C	0.02501 N at 20°C	136

Specification	Reference
Standard Sodium Thiosulfate Solution, 0.0250 N	APHA (4500-S2- F)
Standard Sodium Thiosulfate Titrant	APHA (4500-O D)
Standard Sodium Thiosulfate Titrant	APHA (4500-O E)
Standard Sodium Thiosulfate Titrant	APHA (4500-O F)
Standard Sodium Thiosulfate Titrant, 0.025 N	APHA (4500-CI B)
Standard Sodium Thiosulfate Titrant	APHA (4500-O C)
Standard Sodium Thiosulfate Titrant, 0.025 M	APHA (5530 C)
Standard Sodium Thiosulfate Solution (0.025 N)	EPA (SW-846) (9031)
Standard Sodium Thiosulfate solution (0.025 N)	EPA (SW-846) (9034)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
7900-1	4 L natural poly	18 months
7900-16	500 mL natural poly	18 months
7900-1CT	4 L Cubitainer®	18 months
7900-32	1 L natural poly	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (03/29/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --
Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Certificate of Analysis

Alkaline-Iodide-Azide, Pomeroy Formulation for Dissolved Oxygen (DO) Analysis

Lot Number: 1405D67

Product Number: 535

Manufacture Date: APR 05, 2024

Expiration Date: APR 2026

This solution is intended for use with samples with high Dissolved Oxygen content (above 15 mg/L) and for samples with high concentrations of organic material.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Iodide	7681-82-5	ACS
Sodium Hydroxide	1310-73-2	ACS
Sodium Azide	26628-22-8	Reagent

Test	Specification	Result
Appearance	Colorless liquid	Passed
Free Iodine	To Pass Test	Passed

Specification	Reference
Alkaline Iodide-Sodium Azide Solution II	ASTM (D 888 A)
Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.	

Part Number	Size / Package Type	Shelf Life (Unopened Container)
535-32	1 L natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Heidi J Green (04/05/2024)
Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

W3110
58
operate!
06/27/2024

Certificate of Analysis

1 Reagent Lane
Fair Lawn, NJ 07410
201.796.7100 tel
201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
Standard ISO9001:2015 by SAI Global Certificate Number CERT - 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	02/23/2024
Lot Number	235898		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Feb/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	73
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.670
EVAPORATION RESIDUE	ppm	<= 1	0.3
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.64
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.16
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.06
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.002
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.380
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Harout Sahagian

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.
If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.



Certificate of Analysis



Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Calcium	<= 0.005 %	<0.005 %	PASS
Chloride	<= 0.005 %	0.002 %	PASS
Heavy Metals	<= 0.002 %	<0.002 %	PASS
Iron	<= 0.001 %	<0.001 %	PASS
Magnesium	<= 0.002 %	<0.002 %	PASS
Mercury	<= 0.1 ppm	<0.1 ppm	PASS
Nickel	<= 0.001 %	<0.001 %	PASS
Nitrogen Compounds	<= 0.001 %	<0.001 %	PASS
Phosphate	<= 0.001 %	<0.001 %	PASS
Potassium	<= 0.02 %	<0.02 %	PASS
Purity	>= 97.0 %	99.2 %	PASS
Sodium Carbonate	<= 1.0 %	0.5 %	PASS
Sulfate	<= 0.003 %	<0.003 %	PASS

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.



Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

Spec Set: 0583ACS

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.

Item Number	ED150	Lot Number	2ND0156
Item	Edetate Disodium, Dihydrate, USP	CAS Number	6381-92-6
Molecular Formula	$C_{10}H_{14}N_2Na_2O_8 \cdot 2H_2O$	Molecular Weight	372.24

TEST	SPECIFICATION		RESULT
	MIN	MAX	
ASSAY (DRIED BASIS)	99.0	101.0 %	99.5 %
pH OF A 5% SOLUTION @ 25°C	4.0	6.0	4.6
LOSS ON DRYING	8.7	11.4 %	8.90 %
CALCIUM (Ca)	NO PRECIPITATE IS FORMED		NO PRECIPITATE IS FORMED
ELEMENTAL IMPURITIES:			.
NICKEL (Ni)	AS REPORTED		<0.3 ppm
CHROMIUM (Cr)	AS REPORTED		<0.3 ppm
NITRILOTRIACETIC ACID[n[(HOCOCH ₂) ₃ N]		0.1 %	<0.10 %
IDENTIFICATION A	MATCHES REFERENCE		MATCHES REFERENCE
IDENTIFICATION B	RED COLOR IS DISCHARGED, LEAVING A YELLOWISH SOLUTION		RED COLOR IS DISCHARGED, LEAVING A YELLOWISH SOLUTION
IDENTIFICATION C	MEETS THE REQUIREMENTS FOR SODIUM		MEETS THE REQUIREMENTS FOR SODIUM
CERTIFIED HALAL			CERTIFIED HALAL
EXPIRATION DATE			10-JUL-2026
DATE OF MANUFACTURE			11-JUL-2023
APPEARANCE			WHITE CRYSTALLINE POWDER
RESIDUAL SOLVENTS		AS REPORTED	NO RESIDUAL SOLVENTS PRESENT
MONOGRAPH EDITION			USP 2024

Certificate of Analysis Results Entered By:

CACEVEDO
Charmian Acevedo
22-MAY-24 08:12:30

Certificate of Analysis Results Approved By:

GHERRERA
Genaro Herrera
22-MAY-24 12:32:01

Spectrum Chemical Mfg Corp
755 Jersey Avenue
New Brunswick 08901 NJ



All pharmaceutical ingredients are tested using current edition of applicable pharmacopeia.

Read and understand label and SDS before handling any chemicals. All Spectrum's chemicals are for manufacturing, processing, repacking or research purposes by experienced personnel only. It is the customer's responsibility to provide adequate hazardous material training and ensure that appropriate Personal Protective Equipment (PPE) is used before handling any chemical.

The Elemental Impurities standards implemented by USP and other Pharmaceutical Compendia reflect a growing understanding of the toxicology of trace levels of elemental impurities that can remain in drug substances originating from either raw materials or manufacturing processes. Identifying and quantifying impurities can be critical to predicting the best possible patient outcomes. Elemental Impurities has been a requirement of all products meeting USP/NF, EP and BP monographs since January 1, 2018. More information can be found in USP sections <232> Elemental Impurities – Limits and <233> Elemental Impurities – Procedures. Data for drug substances furnished by Spectrum Chemical Mfg. Corp can be used to ensure that patient daily exposures by oral administration to the selected elements are not exceeded in the formulation of pharmaceutical products.

Certificate of Analysis

Sodium Hypochlorite Solution, 5% available Chlorine

Lot Number: 2407F34**Product Number:** 7495.5**Manufacture Date:** JUL 12, 2024**Expiration Date:** JAN 2025

This solution is subject to slow decomposition upon exposure to air. Keep container tightly capped. Refrigeration may improve stability.
When used in the Phenate method for Ammonia, APHA recommends replacing this solution about every 2 months.

Name	CAS#	Grade
Water	7732-18-5	Commercial
Sodium Hypochlorite	7681-52-9	Commercial

Test	Specification	Result	NIST SRM#
Appearance	Colorless to greenish-yellow liquid	Passed	
Assay (vs. Sodium Thiosulfate/Starch)	4.75-5.25 % (w/w) Cl ₂	5.05 % (w/w) Cl ₂	136

Specification	Reference
Sodium Hypochlorite, 5%	APHA (4500-NH3 F)
Sodium Hypochlorite	ASTM (D 4785)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
7495.5-1	4 L black poly	6 months
7495.5-16	500 mL amber poly	6 months
7495.5-32	1 L amber poly	6 months
7495.5-8	250 mL amber poly	6 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Jose Pena (07/12/2024)
Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



An ISO 9001 Certified Company

Loveland, CO 80539

(970) 669-3050

Certificate of Analysis

This is a Component of 1486266 / LOT A4169

PRODUCT: BOD Nutrient Buffer Pillows

PRODUCT NUMBER: 1486227

LOT NUMBER: A4169

MANUFACTURE DATE: 06/24/2024

DATE OF ANALYSIS: 07/03/2024

TEST	SPECIFICATIONS	RESULTS
Calcium Concentration of a diluted pillow	0.93 to 1.29 ppm	0.960 ppm
Magnesium Concentration of a diluted pillow	0.35 to 0.48 ppm	0.390 ppm
pH in a 6 L of DI water	7.1 to 7.6	7.37
Ammonia Concentration of a diluted pillow	0.57 to 0.79 ppm	0.593 ppm
Iron Concentration of a diluted pillow	0.27 to 0.36 ppm	0.311 ppm
Sterility	To Pass	Passed
Phosphorus Concentration of a diluted pillow	7.6 to 10.3 ppm	8.32 ppm
Five Day Change in Dissolved Oxygen Concentration	-0.2 to 0.2 ppm	0.03 ppm

The expiration date is Jun 2029

Certified by: *Scott Als*

Analytical Services Chemist



Certificate of Analysis

Starch Indicator, 0.5% (w/v), Mercury Free, for Iodometric Titrations

Lot Number: 4408P62

Product Number: 8000

Manufacture Date: AUG 28, 2024

Expiration Date: AUG 2026

This product is Mercury-free.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Starch, soluble	9005-84-9	ACS
Salicylic Acid	69-72-7	ACS

Test	Specification	Result
Appearance	White translucent liquid	Passed
Suitability for Use	Colorless (Iodine absent) - Blue (Iodine present)	Passed

Specification	Reference
Starch Solution	APHA (4500-S2- F)
Starch Indicator Solution	APHA (4500-CI B)
Starch Indicator	APHA (4500-SO32- B)
Starch indicator solution	APHA (2350 B)
Starch indicator solution	APHA (2350 E)
Starch Solution	APHA (510 B)
Starch Solution	APHA (5530 C)
Starch Indicator	APHA (4500-CI C)
Starch Indicator	EPA (345.1)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
8000-1	4 L natural poly	24 months
8000-16	500 mL natural poly	24 months
8000-32	1 L natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

A handwritten signature in blue ink that reads "Paul Brandon". The signature is fluid and cursive, with the first name "Paul" and last name "Brandon" clearly distinguishable.

Paul Brandon (08/28/2024)
Production Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.