

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q1101

Project: Pre Treatment Plant

RunNo.: LB134295

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Orthophosphate as P	mg/L	0.511	0.50	102	90-110	01/15/2025
Sample ID: CCV1 Orthophosphate as P	mg/L	0.509	0.5	102	90-110	01/15/2025
Sample ID: CCV2 Orthophosphate as P	mg/L	0.506	0.5	101	90-110	01/15/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q1101

Project: Pre Treatment Plant

RunNo.: LB134305

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	01/16/2025
Sample ID: CCV1 Ammonia as N	mg/L	1	1	100	90-110	01/16/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	01/16/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	01/16/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q1101

Project: Pre Treatment Plant

RunNo.: LB134307

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.483	0.50	97	90-110	01/16/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.489	0.50	98	90-110	01/16/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.480	0.50	96	90-110	01/16/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q1101

Project: Pre Treatment Plant

RunNo.: LB134307

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------