

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1101

Project: Pre Treatment Plant

Run No.: LB134295

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134295BS							
Orthophosphate as P	mg/L	0.5	0.51		102	1	90-110	01/15/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1101

Project: Pre Treatment Plant

Run No.: LB134299

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134299BS							
Oil and Grease	mg/L	20.0	16.9		84	1	78-114	01/16/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1101

Project: Pre Treatment Plant

Run No.: LB134301

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134301BS							
BOD5	mg/L	198	193		97	1	84.6-115.4	01/16/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1101

Project: Pre Treatment Plant

Run No.: LB134304

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134304BS							
TSS	mg/L	550	510		93	1	90-110	01/16/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1101

Project: Pre Treatment Plant

Run No.: LB134305

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB166064BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	01/16/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1101

Project: Pre Treatment Plant

Run No.: LB134307

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB166065BS							
Phosphorus, Total	mg/L	0.50	0.48		97	1	90-110	01/16/2025