

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51955 MA No.: SDG No.: YE8G3
Instrument ID: CV1 Start Date: 02/05/2025
Analytical Method: CVAA Run Batch: LB134574
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.0	0	0.2	0.19	-3	2.5	2.4	-3

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51955 MA No.: _____ SDG No.: YE8G3
Instrument ID: CV1 Start Date: 02/05/2025
Analytical Method: CVAA Run Batch: LB134574
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	5.0	1	7.5	7.8	4	10	9.8	-2