

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51955 MA No. : SDG No.: YE8G3
 Initial Calibration Verification Source : MP84352
 Continuing Calibration Verification Source : MP84354
 Run Batch: LB134574 Analytical Method: CVAA
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV084				ID: CCV021				ID: CCV022		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	4.0	99		5.0	5.0	100		4.8	96	

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51955 MA No. : SDG No.: YE8G3
Initial Calibration Verification Source : MP84352
Continuing Calibration Verification Source : MP84354
Run Batch: LB134574 Analytical Method: CVAA
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV023				ID: CCV024		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury					5.0	4.8	96		4.8	95	