

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB134440

Review By	Sarabjit Jaswal	Review On	1/28/2025 10:53:54 AM
Supervise By	Mohan Bera	Supervise On	1/28/2025 4:22:54 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84042,MP84050,MP84049,MP84047,MP84046,MP84045,MP84044,MP84043,MP84073		
ICV Standard	MP84051		
CCV Standard	MP84052		
ICSA Standard	MP84053,MP84054		
CRI Standard			
LCS Standard	MP84099		
Chk Standard	MP84057,MP84056		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE001	TUNE	01/24/25 08:56		Jaswal	OK
2	S00	S0	CAL1	01/24/25 09:47		Jaswal	OK
3	S02	S02	CAL3	01/24/25 09:53		Jaswal	OK
4	S03	S03	CAL4	01/24/25 09:57		Jaswal	OK
5	S04	S04	CAL5	01/24/25 10:00		Jaswal	OK
6	S05	S05	CAL6	01/24/25 10:03		Jaswal	OK
7	S06	S06	CAL7	01/24/25 10:06		Jaswal	OK
8	S07	S07	CAL8	01/24/25 10:09		Jaswal	OK
9	S08	S08	CAL9	01/24/25 10:11		Jaswal	OK
10	ICV001	ICV001	ICV	01/24/25 10:23	Inst. paused for calibration review	Jaswal	OK
11	ICB001	ICB001	ICB	01/24/25 10:30	Inst. paused for review	Jaswal	OK
12	ICSA001	ICSA001	ICSA	01/24/25 10:33		Jaswal	OK
13	ICSAB001	ICSAB001	ICSAB	01/24/25 10:36		Jaswal	OK
14	CCV001	CCV001	CCV	01/24/25 10:41		Jaswal	OK
15	CCB001	CCB001	CCB	01/24/25 10:45		Jaswal	OK
16	P4368-01	LOD-MDL-SOIL-01-Q	SAM	01/24/25 11:01	Not Use	Jaswal	Not Ok
17	P4368-02	LOQ-SOIL-02-QT4-20	SAM	01/24/25 11:13	Not Use	Jaswal	Not Ok
18	P4368-16	LLOQ-SOIL-QT4-2024	SAM	01/24/25 11:17	Not Use	Jaswal	Not Ok

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB134440

Review By	Sarabjit Jaswal	Review On	1/28/2025 10:53:54 AM
Supervise By	Mohan Bera	Supervise On	1/28/2025 4:22:54 PM

STD. NAME	STD REF.#
ICAL Standard	MP84042,MP84050,MP84049,MP84047,MP84046,MP84045,MP84044,MP84043,MP84073
ICV Standard	MP84051
CCV Standard	MP84052
ICSA Standard	MP84053,MP84054
CRI Standard	
LCS Standard	MP84099
Chk Standard	MP84057,MP84056

19	P4368-07	LOD-MDL-WATER-01	SAM	01/24/25 11:20	Not Use	Jaswal	Not Ok
20	P4368-08	LOQ-WATER-02-QT4	SAM	01/24/25 11:26	Not Use	Jaswal	Not Ok
21	P4368-17	LLOQ-WATER-QT4-2	SAM	01/24/25 11:29	Not Use	Jaswal	Not Ok
22	P4368-10	LOD-MDL-WATER-04	SAM	01/24/25 11:35	Not Use	Jaswal	Not Ok
23	P4368-11	LOQ-WATER-05-QT4	SAM	01/24/25 11:38	Not Use	Jaswal	Not Ok
24	CCV002	CCV002	CCV	01/24/25 11:49		Jaswal	OK
25	CCB002	CCB002	CCB	01/24/25 11:52		Jaswal	OK
26	PBW	PBW	SAM	01/24/25 11:59	Not Use	Jaswal	Not Ok
27	P4476-03	XA102	SAM	01/24/25 12:02	Not Use	Jaswal	Not Ok
28	P4476-01	XA100	SAM	01/24/25 12:06	Not Use	Jaswal	Not Ok
29	LCSW	LCSW	SAM	01/24/25 12:22	Not Use	Jaswal	Not Ok
30	PBS	PBS	LCS	01/24/25 12:26	Not Use	Jaswal	Not Ok
31	LCSS	LCSS	SAM	01/24/25 12:29	Not Use	Jaswal	Not Ok
32	PBW 6020	PBW 6020	SAM	01/24/25 12:33	Not Use	Jaswal	Not Ok
33	LCSW 6020	LCSW 6020	SAM	01/24/25 12:36	Not Use	Jaswal	Not Ok
34	PBS 6020	PBS 6020	SAM	01/24/25 12:42	Not Use	Jaswal	Not Ok
35	LCSS 6020	LCSS 6020	SAM	01/24/25 12:46	Not Use	Jaswal	Not Ok
36	CCV003	CCV003	CCV	01/24/25 12:49		Jaswal	OK
37	CCB003	CCB003	CCB	01/24/25 12:56	Inst. paused for review	Jaswal	OK
38	PBW 200.8	PBW 200.8	SAM	01/24/25 12:59	Not Use	Jaswal	Not Ok

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB134440

Review By	Sarabjit Jaswal	Review On	1/28/2025 10:53:54 AM
Supervise By	Mohan Bera	Supervise On	1/28/2025 4:22:54 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84042,MP84050,MP84049,MP84047,MP84046,MP84045,MP84044,MP84043,MP84073		
ICV Standard	MP84051		
CCV Standard	MP84052		
ICSA Standard	MP84053,MP84054		
CRI Standard			
LCS Standard	MP84099		
Chk Standard	MP84057,MP84056		

39	LCSW 200.8	LCSW 200.8	SAM	01/24/25 13:03	Not Use	Jaswal	Not Ok
40	CCV004	CCV004	CCV	01/24/25 13:05		Jaswal	OK
41	CCB004	CCB004	CCB	01/24/25 13:19	Inst. paused for review	Jaswal	OK
42	PB166174BL	PBW174	MB	01/24/25 13:22		Jaswal	OK
43	PB166174BS	LCS174	LCS	01/24/25 13:25		Jaswal	OK
44	Q1089-01	MC0B00	SAM	01/24/25 13:29		Jaswal	OK
45	Q1089-02	MC0B01	SAM	01/24/25 13:32		Jaswal	OK
46	Q1089-03	MC0B01D	DUP	01/24/25 13:35		Jaswal	OK
47	Q1089-02L	MC0B01L	SD	01/24/25 13:39		Jaswal	OK
48	Q1089-04	MC0B01S	MS	01/24/25 13:42		Jaswal	OK
49	Q1089-05	MC0B06	SAM	01/24/25 13:45		Jaswal	OK
50	Q1089-06	MC0B07	SAM	01/24/25 13:49		Jaswal	OK
51	Q1089-07	MC0AL9	SAM	01/24/25 13:52		Jaswal	OK
52	Q1089-08	MC0AM2	SAM	01/24/25 13:55		Jaswal	OK
53	Q1089-09	MC0AZ9	SAM	01/24/25 13:59		Jaswal	OK
54	Q1089-10	MC0B02	SAM	01/24/25 14:02		Jaswal	OK
55	Q1089-11	MC0B04	SAM	01/24/25 14:05		Jaswal	OK
56	Q1089-12	MC0B09	SAM	01/24/25 14:08		Jaswal	OK
57	Q1089-13	MC0B10	SAM	01/24/25 14:12		Jaswal	OK
58	Q1089-14	MC0B11	SAM	01/24/25 14:15		Jaswal	OK

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB134440

Review By	Sarabjit Jaswal	Review On	1/28/2025 10:53:54 AM
Supervise By	Mohan Bera	Supervise On	1/28/2025 4:22:54 PM

STD. NAME	STD REF.#
ICAL Standard	MP84042,MP84050,MP84049,MP84047,MP84046,MP84045,MP84044,MP84043,MP84073
ICV Standard	MP84051
CCV Standard	MP84052
ICSA Standard	MP84053,MP84054
CRI Standard	
LCS Standard	MP84099
Chk Standard	MP84057,MP84056

59	PB166148BL	PBS148	MB	01/24/25 14:18		Jaswal	OK
60	PB166148BS	LCS148	LCS	01/24/25 14:24		Jaswal	OK
61	Q1102-01	MH2H28	SAM	01/24/25 14:27		Jaswal	OK
62	CCV005	CCV005	CCV	01/24/25 14:31		Jaswal	OK
63	CCB005	CCB005	CCB	01/24/25 14:37	Inst. paused for review	Jaswal	OK
64	Q1102-02	MH2H29	SAM	01/24/25 14:40		Jaswal	OK
65	Q1102-03	MH2H30	SAM	01/24/25 14:44		Jaswal	OK
66	Q1102-04	MH2H30D	DUP	01/24/25 14:47		Jaswal	OK
67	Q1102-03L	MH2H30L	SD	01/24/25 14:50		Jaswal	OK
68	Q1102-05	MH2H30S	MS	01/24/25 14:53	MS Fail For As,Cu	Jaswal	OK
69	Q1102-06	MH2H31	SAM	01/24/25 14:57		Jaswal	OK
70	PB166154BL	PBS154	MB	01/24/25 15:03		Jaswal	OK
71	PB166154BS	LCS154	LCS	01/24/25 15:33		Jaswal	OK
72	Q1105-01	YE8G3	SAM	01/24/25 15:36		Jaswal	OK
73	Q1105-02	YE8G5	SAM	01/24/25 15:39		Jaswal	OK
74	Q1105-03	YE8G6	SAM	01/24/25 15:42	Sr High in MS	Jaswal	Dilution
75	Q1105-04	YE8G6D	DUP	01/24/25 15:46	Sr High in MS	Jaswal	Dilution
76	Q1105-03L	YE8G6L	SD	01/24/25 15:49	Sr High in MS	Jaswal	Dilution
77	Q1105-05	YE8G6S	MS	01/24/25 15:52	Sr High , MS Fail For Mn	Jaswal	Dilution
78	Q1105-06	YE8H0	SAM	01/24/25 15:56		Jaswal	OK

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB134440

Review By	Sarabjit Jaswal	Review On	1/28/2025 10:53:54 AM
Supervise By	Mohan Bera	Supervise On	1/28/2025 4:22:54 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84042,MP84050,MP84049,MP84047,MP84046,MP84045,MP84044,MP84043,MP84073		
ICV Standard	MP84051		
CCV Standard	MP84052		
ICSA Standard	MP84053,MP84054		
CRI Standard			
LCS Standard	MP84099		
Chk Standard	MP84057,MP84056		

79	Q1105-07	YE8H1	SAM	01/24/25 15:59		Jaswal	OK
80	Q1105-08	YE8H2	SAM	01/24/25 16:02	As High	Jaswal	Dilution
81	Q1105-09	YE8H3	SAM	01/24/25 16:05		Jaswal	OK
82	Q1105-10	YE8H4	SAM	01/24/25 16:09		Jaswal	OK
83	CCV006	CCV006	CCV	01/24/25 16:12		Jaswal	OK
84	CCB006	CCB006	CCB	01/24/25 16:18	Inst. paused for review	Jaswal	OK
85	Q1105-11	YE8H5	SAM	01/24/25 16:22		Jaswal	OK
86	Q1105-12	YE8H6	SAM	01/24/25 16:25		Jaswal	OK
87	Q1105-13	YE8E9	SAM	01/24/25 16:28		Jaswal	OK
88	Q1105-14	YE8F0	SAM	01/24/25 16:31		Jaswal	OK
89	Q1105-15	YE8F1	SAM	01/24/25 16:35		Jaswal	OK
90	Q1105-16	YE8F2	SAM	01/24/25 16:38		Jaswal	OK
91	Q1105-17	YE8F3	SAM	01/24/25 16:41		Jaswal	OK
92	Q1105-18	YE8F4	SAM	01/24/25 16:45		Jaswal	OK
93	Q1105-19	YE8F6	SAM	01/24/25 16:48		Jaswal	OK
94	Q1105-20	YE8G0	SAM	01/24/25 16:51		Jaswal	OK
95	Q1105-21	YE8G1	SAM	01/24/25 16:54		Jaswal	OK
96	Q1105-22	YE8G2	SAM	01/24/25 16:58		Jaswal	OK
97	PB166195BL	PBS195	MB	01/24/25 17:01		Jaswal	OK
98	PB166195BS	LCS195	LCS	01/24/25 17:11		Jaswal	OK

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB134440

Review By	Sarabjit Jaswal	Review On	1/28/2025 10:53:54 AM
Supervise By	Mohan Bera	Supervise On	1/28/2025 4:22:54 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84042,MP84050,MP84049,MP84047,MP84046,MP84045,MP84044,MP84043,MP84073		
ICV Standard	MP84051		
CCV Standard	MP84052		
ICSA Standard	MP84053,MP84054		
CRI Standard			
LCS Standard	MP84099		
Chk Standard	MP84057,MP84056		

99	Q1135-01	YE8F5	SAM	01/24/25 17:14		Jaswal	OK
100	Q1135-02	YE8F7	SAM	01/24/25 17:17	Sr High in MS	Jaswal	Dilution
101	Q1135-03	YE8F7D	DUP	01/24/25 17:21	Sr High in MS	Jaswal	Dilution
102	Q1135-02L	YE8F7L	SD	01/24/25 17:24	Sr High in MS	Jaswal	Dilution
103	Q1135-04	YE8F7S	MS	01/24/25 17:27	Sr High,MS Fail For MO(4.33ppb)	Jaswal	Dilution
104	Q1135-05	YE8F8	SAM	01/24/25 17:30		Jaswal	OK
105	Q1135-06	YE8F9	SAM	01/24/25 17:34		Jaswal	OK
106	Q1135-07	YE8H9	SAM	01/24/25 17:37		Jaswal	OK
107	Q1135-08	YE8J0	SAM	01/24/25 17:40		Jaswal	OK
108	Q1135-10	YE8J1	SAM	01/24/25 17:43		Jaswal	OK
109	Q1135-11	YE8E3	SAM	01/24/25 17:47		Jaswal	OK
110	Q1135-12	YE8E4	SAM	01/24/25 17:50		Jaswal	OK
111	CCV007	CCV007	CCV	01/24/25 17:58		Jaswal	OK
112	CCB007	CCB007	CCB	01/24/25 18:01		Jaswal	OK
113	Q1135-13	YE8E5	SAM	01/24/25 18:04		Jaswal	OK
114	Q1135-14	YE8E6	SAM	01/24/25 18:08		Jaswal	OK
115	Q1135-15	YE8E7	SAM	01/24/25 18:12		Jaswal	OK
116	Q1135-16	YE8E8	SAM	01/24/25 18:16		Jaswal	OK
117	Q1135-17	YE8D2	SAM	01/24/25 18:19		Jaswal	OK

Instrument ID: P8

Daily Analysis Runlog For Sequence/QC Batch ID # LB134440

Review By	Sarabjit Jaswal	Review On	1/28/2025 10:53:54 AM
Supervise By	Mohan Bera	Supervise On	1/28/2025 4:22:54 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84042,MP84050,MP84049,MP84047,MP84046,MP84045,MP84044,MP84043,MP84073		
ICV Standard	MP84051		
CCV Standard	MP84052		
ICSA Standard	MP84053,MP84054		
CRI Standard			
LCS Standard	MP84099		
Chk Standard	MP84057,MP84056		

118	Q1135-18	YE8D4	SAM	01/24/25 18:23		Jaswal	OK
119	Q1135-19	YE8C3	SAM	01/24/25 18:26		Jaswal	OK
120	Q1135-21	YE8D0	SAM	01/24/25 18:29		Jaswal	OK
121	Q1135-22	YE8D5	SAM	01/24/25 18:33		Jaswal	OK
122	Q1105-03DL	YE8G6	SAM	01/24/25 18:36	5X For Sr	Jaswal	Confirms
123	Q1105-04DL	YE8G6D	DUP	01/24/25 18:39	5X For Sr	Jaswal	Confirms
124	Q1105-03LDL	YE8G6L	SD	01/24/25 18:43	5X For Sr	Jaswal	Confirms
125	Q1105-05DL	YE8G6S	MS	01/24/25 18:46	5X For Sr	Jaswal	Confirms
126	Q1105-08DL	YE8H2	SAM	01/24/25 18:49	2X For As	Jaswal	Confirms
127	Q1135-02DL	YE8F7	SAM	01/24/25 18:53	5X For Sr	Jaswal	Confirms
128	Q1135-03DL	YE8F7D	DUP	01/24/25 18:56	5X For Sr	Jaswal	Confirms
129	Q1135-02LDL	YE8F7L	SD	01/24/25 18:59	5X For Sr	Jaswal	Confirms
130	Q1135-04DL	YE8F7S	MS	01/24/25 19:03	5X For Sr	Jaswal	Confirms
131	Q1102-03A	MH2H30A	PS	01/24/25 19:06	PS For As,Cu(0.0027ml mp84220,0.0023ml m5751-10ml sample)	Jaswal	OK
132	Q1105-03A	YE8G6A	PS	01/24/25 19:09	PS For Mn(0.005ml m6128-10ml sample)	Jaswal	OK
133	CCV008	CCV008	CCV	01/24/25 19:19		Jaswal	OK
134	CCB008	CCB008	CCB	01/24/25 19:22		Jaswal	OK