

YE8G6A

POST-DIGESTION/DISTILLATION SPIKE SAMPLE RECOVERY

Lab Name:	<u>Alliance Technical Group, LLC</u>	Contract:	<u>68HERH20D0011</u>
Lab Code:	<u>ACE</u>	Case No.:	<u>51955</u>
		MA No. :	<u> </u>
		SDG No.:	<u>YE8G5</u>
Matrix :	<u>Water</u>	Analytical Method:	<u>ICP-AES</u>
% Solids:	<u> </u>		
Concentration Units (µg/L, mg/L or mg/kg dry weight):		<u> ug/L</u>	

Analyte	Control Limit %R	Spiked Sample Result (SSR) Q		Sample Result (SR) Q		Spike Added (SA)	%R	Q
Aluminum	75 - 125	1400		370		740	139	
Manganese	75 - 125	35		3.4	J	30.0	105	