


28134860

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.3011e-004 slope

C= -5.3637e-002 y-intercept

Rho= 0.9993159

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.042	-0.042	50	0.000	50					
0.05	0.050										
0.20	0.200	0.164	-0.036	946	0.0 %	946					
2.50	2.500	2.470	-0.030	10966	0.0 %	10966					
5.00	5.000	5.293	0.293	23236	0.0 %	23236					
7.50	7.500	7.359	-0.141	32213	0.0 %	32213					
10.0	10.000	9.956	-0.044	43502	0.0 %	43502					

0/010
= 18
- 1
6
- 2
0

LB134860

INSTRUMENT ID: CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	50	-		0 SFAM01.1	PPB	02/28/2025 13:43	Std
	0.2 S01	946	-		0.2 SFAM01.1	PPB	02/28/2025 13:45	Std
	2.5 S02	10966	-		2.5 SFAM01.1	PPB	02/28/2025 13:47	Std
	5 S03	23236	-		5 SFAM01.1	PPB	02/28/2025 13:50	Std
	7.5 S04	32213	-		7.5 SFAM01.1	PPB	02/28/2025 13:52	Std
	10 S05	43502	-		10 SFAM01.1	PPB	02/28/2025 13:54	Std
ICV024	ICV024	17588	3.9935 -		SFAM01.1	PPB	02/28/2025 13:57	SMPL
ICB024	ICB024	-251	-0.1114 -		SFAM01.1	PPB	02/28/2025 13:59	SMPL
CCV045	CCV045	23261	5.2989 -		SFAM01.1	PPB	02/28/2025 14:02	SMPL
CCB045	CCB045	-246	-0.1102 -		SFAM01.1	PPB	02/28/2025 14:04	SMPL
PB166941BL	PBW941	55	-0.041 -		SFAM01.1	PPB	02/28/2025 14:06	SMPL
Q1106-01	YE8G3	70	-0.0375 -		SFAM01.1	PPB	02/28/2025 14:08	SMPL
Q1106-02	YE8G5	169	-0.0147 -		SFAM01.1	PPB	02/28/2025 14:11	SMPL
Q1106-03	YE8G6	304	0.0163 -		SFAM01.1	PPB	02/28/2025 14:13	SMPL
Q1106-04	YE8G6D	363	0.0299 -		SFAM01.1	PPB	02/28/2025 14:15	SMPL
Q1106-05	YE8G6S	5135	1.128 -		SFAM01.1	PPB	02/28/2025 14:17	SMPL
Q1106-06	YE8H0	416	0.0421 -		SFAM01.1	PPB	02/28/2025 14:20	SMPL
Q1106-07	YE8H1	554	0.0738 -		SFAM01.1	PPB	02/28/2025 14:22	SMPL
Q1106-08	YE8H2	922	0.1585 -		SFAM01.1	PPB	02/28/2025 14:24	SMPL
Q1106-09	YE8H3	353	0.0276 -		SFAM01.1	PPB	02/28/2025 14:27	SMPL
Q1106-10	YE8H4	694	0.1061 -		SFAM01.1	PPB	02/28/2025 14:29	SMPL
Q1106-11	YE8H5	463	0.0529 -		SFAM01.1	PPB	02/28/2025 14:31	SMPL
Q1106-12	YE8H6	274	0.0094 -		SFAM01.1	PPB	02/28/2025 14:33	SMPL
Q1106-13	YE8F6	353	0.0276 -		SFAM01.1	PPB	02/28/2025 14:36	SMPL
Q1106-14	YE8F9	299	0.0152 -		SFAM01.1	PPB	02/28/2025 14:38	SMPL
Q1106-15	YE8G0	357	0.0285 -		SFAM01.1	PPB	02/28/2025 14:40	SMPL
Q1106-16	YE8G1	339	0.0244 -		SFAM01.1	PPB	02/28/2025 14:42	SMPL
Q1106-17	YE8G2	272	0.009 -		SFAM01.1	PPB	02/28/2025 14:45	SMPL
Q1106-18	YE8H9	271	0.0087 -		SFAM01.1	PPB	02/28/2025 14:47	SMPL
Q1106-19	YE8J0	454	0.0508 -		SFAM01.1	PPB	02/28/2025 14:49	SMPL
CCV046	CCV046	23580	5.3723 -		SFAM01.1	PPB	02/28/2025 14:51	SMPL
CCB046	CCB046	-258	-0.113 -		SFAM01.1	PPB	02/28/2025 14:54	SMPL
Q1106-20	YE8J1	206	-0.0062 -		SFAM01.1	PPB	02/28/2025 14:56	SMPL
PB166943BL	PBW943	70	-0.0375 -		SFAM01.1	PPB	02/28/2025 14:58	SMPL
Q1160-01	YE8C9	239	0.0014 -		SFAM01.1	PPB	02/28/2025 15:01	SMPL
Q1160-02	YE8C9D	201	-0.0074 -		SFAM01.1	PPB	02/28/2025 15:03	SMPL
Q1160-03	YE8C9S	5373	1.1827 -		SFAM01.1	PPB	02/28/2025 15:05	SMPL
Q1160-04	YE8D0	338	0.0241 -		SFAM01.1	PPB	02/28/2025 15:07	SMPL
Q1160-05	YE8D3	304	0.0163 -		SFAM01.1	PPB	02/28/2025 15:10	SMPL
PB166945BL	PBW945	97	-0.0313 -		SFAM01.1	PPB	02/28/2025 15:12	SMPL
Q1136-01	YE8F5	82	-0.0348 -		SFAM01.1	PPB	02/28/2025 15:14	SMPL
Q1136-02	YE8F7	125	-0.0249 -		SFAM01.1	PPB	02/28/2025 15:16	SMPL
Q1136-03	YE8F7D	88	-0.0334 -		SFAM01.1	PPB	02/28/2025 15:19	SMPL
Q1136-04	YE8F7S	5497	1.2113 -		SFAM01.1	PPB	02/28/2025 15:21	SMPL
Q1136-05	YE8F8	-20	-0.0582 -		SFAM01.1	PPB	02/28/2025 15:23	SMPL
Q1136-06	YE8E3	41	-0.0442 -		SFAM01.1	PPB	02/28/2025 15:26	SMPL

LB134860

INSTRUMENT ID: CV1

Q1136-07	YE8E4	112	-0.0279 -	SFAM01.1 PPB	02/28/2025 15:28 SMPL
Q1136-08	YE8E5	423	0.0437 -	SFAM01.1 PPB	02/28/2025 15:30 SMPL
Q1136-09	YE8E6	4	-0.0527 -	SFAM01.1 PPB	02/28/2025 15:32 SMPL
Q1136-10	YE8E7	63	-0.0391 -	SFAM01.1 PPB	02/28/2025 15:35 SMPL
Q1136-11	YE8E8	92	-0.0325 -	SFAM01.1 PPB	02/28/2025 15:37 SMPL
Q1136-12	YE8D1	116	-0.0269 -	SFAM01.1 PPB	02/28/2025 15:39 SMPL
CCV047	CCV047	23048	5.2499 -	SFAM01.1 PPB	02/28/2025 15:41 SMPL
CCB047	CCB047	-259	-0.1132 -	SFAM01.1 PPB	02/28/2025 15:44 SMPL
Q1136-13	YE8D2	-5	-0.0548 -	SFAM01.1 PPB	02/28/2025 15:46 SMPL
Q1136-14	YE8D4	137	-0.0221 -	SFAM01.1 PPB	02/28/2025 15:48 SMPL
Q1136-15	YE8C3	64	-0.0389 -	SFAM01.1 PPB	02/28/2025 15:51 SMPL
Q1136-17	YE8D5	79	-0.0355 -	SFAM01.1 PPB	02/28/2025 15:53 SMPL
PB166941TB	LEB941	107	-0.029 -	SFAM01.1 PPB	02/28/2025 15:55 SMPL
PB166943TB	LEB943	135	-0.0226 -	SFAM01.1 PPB	02/28/2025 15:57 SMPL
PB166945TB	LEB945	109	-0.0286 -	SFAM01.1 PPB	02/28/2025 16:00 SMPL
CCV048	CCV048	22298	5.0773 -	SFAM01.1 PPB	02/28/2025 16:02 SMPL
CCB048	CCB048	-246	-0.1102 -	SFAM01.1 PPB	02/28/2025 16:04 SMPL