



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Weston Solutions

Project Name: Ft Meade Tipton Airfield Parcel RI - PO 0111169

Project # N/A

Chemtech Project # Q1109

Test Name: Diesel Range Organics

A. Number of Samples and Date of Receipt:

3 Water samples were received on 01/16/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Ammonia, Anions Group5, Diesel Range Organics, Gasoline Range Organics, Hardness, Total, Hexavalent Chromium, Mercury, Metals ICP-TAL, METALS-TAL, Oil and Grease, PESTICIDE Group1, PESTICIDE Group3, SVOC-TCL BNA -20, TOC and VOC-TCLVOA-10. This data package contains results for Diesel Range Organics.

C. Analytical Techniques:

The analysis of Diesel Range Organics was based on method 8015D and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

E. Additional Comments:

The not QT review data is reported in the Miscellaneous

F. Calculation for Concentration in WATER samples :

The sample concentrations (Cs) in ug/L are calculated as follows:

$$Cs = \frac{\{\text{Extract DRO-net (ug/mL)}\} \{\text{Final vol. extract (mL)}\} \{\text{Df}\}}{Ws}$$

Where

DRO (net)ug/mL = DRO (total) ug /mL - DRO (solvent) ug /mL

Df = Dilution factor

Ws= Weight of sample in mL

G. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____