



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

### **Weston Solutions**

**Project Name: Ft Meade Tipton Airfield Parcel RI - PO 0111169**

**Project # N/A**

**Chemtech Project # Q1109**

**Test Name: PESTICIDE Group1**

### **A. Number of Samples and Date of Receipt:**

3 Water samples were received on 01/16/2025.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: Ammonia, Anions Group5, Diesel Range Organics, Gasoline Range Organics, Hardness, Total, Hexavalent Chromium, Mercury, Metals ICP-TAL, METALS-TAL, Oil and Grease, PESTICIDE Group1, PESTICIDE Group3, SVOC-TCL BNA -20, TOC and VOC-TCLVOA-10. This data package contains results for PESTICIDE Group1.

### **C. Analytical Techniques:**

The analysis was performed on instrument ECD\_L. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df.; Catalog # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 7HMG017- 11. The analysis of PESTICIDE Group1s was based on method 8081B and extraction was done based on method 3510.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

### **E. Additional Comments:**

The not QT review data is reported in the Miscellaneous.

**F. Calculation for Concentration in Water Samples:**

$$\text{Concentration in ug/L} = \frac{(A_x) (V_t) (DF) (GPC)}{(CF) (V_o) (V_i)}$$

Where,

A<sub>x</sub> = Response (peak area or height) of the compound to be measured.

CF = Mean Calibration Factor from the initial calibration (area/ng).

V<sub>o</sub> = Volume of water extracted in mL.

V<sub>i</sub> = Volume of extract injected in uL.

V<sub>t</sub> = Volume of the concentrated extract in uL

GPC =  $\frac{V_{in}}{V_{out}}$  = GPC factor (If no GPC is performed, GPC=1)

V<sub>in</sub> = Volume of extract loaded onto GPC column.

V<sub>out</sub> = Volume of extract collected after GPC cleanup.

DF = Dilution Factor.

**G. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_