

Duplicate Sample Summary

Client: Weston Solutions	SDG No.: Q1109
Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169	Sample ID: LB134347BS
Client ID: LB134347BSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	16.9		17.1		1	1.18		01/20/2025

Duplicate Sample Summary

Client: Weston Solutions	SDG No.: Q1109
Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169	Sample ID: Q1109-01
Client ID: TAPIAL1-MW04I-011525-00-T3MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TOC	mg/L	+/-20	10.5		10.6		1	1		01/17/2025

Duplicate Sample Summary

Client:	Weston Solutions	SDG No.:	Q1109
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Sample ID:	Q1109-02
Client ID:	TAPIAL1-MW04S-011525-00-T2DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.0030	U	0.0030	U	1	0		01/16/2025

Duplicate Sample Summary

Client: Weston Solutions	SDG No.: Q1109
Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169	Sample ID: Q1109-02
Client ID: TAPIAL1-MW04S-011525-00-T2MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Hexavalent Chromium	mg/L	+/-20	0.98		0.98		2	0.41		01/16/2025
Nitrate	mg/L	+/-15	3.30		3.70		1	11		01/16/2025
Orthophosphate as P	mg/L	+/-15	5.20		5.80		1	11		01/16/2025
Bromide	mg/L	+/-15	9.80		11.1		1	12		01/16/2025
Nitrite	mg/L	+/-15	2.90		3.30		1	13		01/16/2025
Fluoride	mg/L	+/-15	2.00		2.30		1	14		01/16/2025
Chloride	mg/L	+/-15	23.6	OR	24.1	OR	1	2		01/16/2025
Sulfate	mg/L	+/-15	60.3	OR	62.2	OR	1	3		01/16/2025

Duplicate Sample Summary

Client: Weston Solutions	SDG No.: Q1109
Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169	Sample ID: Q1113-01
Client ID: DSN002DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	3.40	OR	3.30	OR	1	3		01/17/2025

Duplicate Sample Summary

Client: Weston Solutions	SDG No.: Q1109
Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169	Sample ID: Q1113-01
Client ID: DSN002MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	4.30	OR	4.10	OR	1	5		01/17/2025