

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Weston Solutions</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q1109</u>
Contract:	<u>WEST04</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1109</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q1092-01</u>	Client ID:	<u>427DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q1092-01DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.30		0.36		20		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Weston Solutions</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q1109</u>
Contract:	<u>WEST04</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1109</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q1092-01MS</u>	Client ID:	<u>427MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q1092-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	3.30		3.04		8		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Weston Solutions **Level:** LOW **SDG No.:** Q1109
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1109 **SAS No.:** Q1109
Matrix: Water **Sample ID:** Q1109-02 **Client ID:** TAPIAL1-MW04S-011525-00-T2DUP
Percent Solids for Sample: NA **Duplicate ID** Q1109-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	453		468		3		P
Antimony	ug/L	20	2.00	U	0.12	J	200.0		P
Arsenic	ug/L	20	0.75	J	0.84	J	11		P
Barium	ug/L	20	27.7		28.0		1		P
Beryllium	ug/L	20	0.19	J	0.17	J	11		P
Cadmium	ug/L	20	1.00	U	1.00	U			P
Calcium	ug/L	20	11200		11400		2		P
Chromium	ug/L	20	1.58	J	1.65	J	4		P
Cobalt	ug/L	20	21.4		21.6		1		P
Copper	ug/L	20	5.61		5.53		1		P
Iron	ug/L	20	665		681		2		P
Lead	ug/L	20	0.89	J	0.85	J	5		P
Magnesium	ug/L	20	3650		3680		1		P
Manganese	ug/L	20	313		313		0		P
Nickel	ug/L	20	13.1		13.1		0		P
Potassium	ug/L	20	977		983		1		P
Selenium	ug/L	20	5.00	U	5.00	U			P
Silver	ug/L	20	1.00	U	1.00	U			P
Sodium	ug/L	20	15200		15100		1		P
Thallium	ug/L	20	0.090	J	1.00	U	200.0		P
Vanadium	ug/L	20	2.50	J	2.60	J	4		P
Zinc	ug/L	20	5.94		6.21		4		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Weston Solutions **Level:** LOW **SDG No.:** Q1109
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1109 **SAS No.:** Q1109
Matrix: Water **Sample ID:** Q1109-02MS **Client ID:** TAPIAL1-MW04S-011525-00-T2MSD
Percent Solids for Sample: NA **Duplicate ID** Q1109-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	8230		8380		2		P
Antimony	ug/L	20	437		443		1		P
Arsenic	ug/L	20	1.05		1.12		6		P
Barium	ug/L	20	2020		2070		2		P
Beryllium	ug/L	20	412		426		3		P
Cadmium	ug/L	20	413		418		1		P
Calcium	ug/L	20	53100		52600		1		P
Chromium	ug/L	20	410		415		1		P
Cobalt	ug/L	20	465		472		1		P
Copper	ug/L	20	4640		4670		1		P
Iron	ug/L	20	20500		20600		0		P
Lead	ug/L	20	2250		2250		0		P
Magnesium	ug/L	20	46700		47100		1		P
Manganese	ug/L	20	4930		4910		0		P
Nickel	ug/L	20	457		461		1		P
Potassium	ug/L	20	20800		21200		2		P
Selenium	ug/L	20	434		435		0		P
Silver	ug/L	20	290		210		32	*	P
Sodium	ug/L	20	56400		57400		2		P
Thallium	ug/L	20	426		430		1		P
Vanadium	ug/L	20	429		436		2		P
Zinc	ug/L	20	4500		4530		1		P