

Analytical Sequence

Client: Weston Solutions

SDG No.: Q1109

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

Instrument ID: FID_F

GC Column: RXI-1MS **ID:** 0.18 (mm)

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

MEAN SUROGATE RT FROM INITIAL CALIBRATION 15.0138					
EPA SAMPLE NO.	LAB SAMPLE ID	DATE AND TIME ANALYZED	DATAFILE	RT	#
PIBLK01	LBLK01	17 Jan 2025 11:17	FF015247.D	15.016	
50 PPM TRPH STD	50 PPM TRPH STD	17 Jan 2025 11:45	FF015248.D	15.018	
TAPIAL1-MW04S-011525-00-T2	Q1109-02	17 Jan 2025 13:39	FF015250.D	15.014	
PB166100BL	PB166100BL	17 Jan 2025 14:07	FF015251.D	15.015	
PB166100BS	PB166100BS	17 Jan 2025 14:36	FF015252.D	15.014	
PB166100BSD	PB166100BSD	17 Jan 2025 15:04	FF015253.D	15.014	
PIBLK02	LBLK02	17 Jan 2025 15:32	FF015254.D	15.015	
50 PPM TRPH STD	50 PPM TRPH STD	17 Jan 2025 16:01	FF015255.D	15.018	

Column used to flag RT values with an * values outside of QC limits

QC Limits
(± 0.10 minutes)

Lower Limit
14.9138

Upper Limits
15.1138