

WATER DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE

Lab Name: Chemtech **Client:** Weston Solutions
Lab Code: CHEM **Cas No:** Q1109 **SAS No :** Q1109 **SDG No:** Q1109
Matrix Spike - EPA Sample No : PB166100BS **Datafile:** FF015252.D

COMPOUND	SPIKE ADDED ug/L	CONCENTRATION ug/L	LCS/LCSD CONCENTRATION ug/L	% REC	QC LIMITS
DRO	200	0	198	99	78-117