



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**WATER DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE**

**Lab Name:** Chemtech **Client:** Weston Solutions  
**Lab Code:** CHEM **Cas No:** Q1109 **SAS No :** Q1109 **SDG No:** Q1109  
**Matrix Spike - EPA Sample No :** PB166100BSD **Datafile:** FF015253.D

COMPOUND	SPIKE ADDED ug/L	CONCENTRATION ug/L	LCS/LCSD CONCENTRATION ug/L	% REC	QC LIMITS
DRO	200	0	197	98	78-117

LCS/LCSD % Recovery RPD : 0.5