

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: WEST04
Lab Code: CHEM Case No.: Q1109 SAS No.: Q1109 SDG No.: Q1109
Instrument ID: BNA_F Calibration Date(s): 01/10/2025 01/10/2025
Calibration Time(s): 11:58 17:45

LAB FILE ID:		RRF2.5 = BF141109.D			RRF005 = BF141110.D		RRF010 = BF141111.D	
		RRF020 = BF141112.D			RRF040 = BF141113.D		RRF050 = BF141117.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.453	1.400	1.378	1.248	1.236	1.295	9.1
Benzaldehyde		1.215	1.139	1.040	0.849	0.782	0.966	19.8
Phenol-d6		1.834	1.773	1.736	1.566	1.569	1.640	8.6
Phenol		1.943	1.902	1.861	1.701	1.665	1.756	8.4
bis(2-Chloroethyl) ether		1.445	1.372	1.363	1.256	1.279	1.308	6.9
2-Chlorophenol		1.561	1.476	1.458	1.345	1.338	1.389	8.2
2-Methylphenol		1.209	1.175	1.182	1.088	1.091	1.121	6.0
2,2-oxybis(1-Chloropropane)		2.046	1.913	1.939	1.766	1.746	1.826	7.8
Acetophenone		0.539	0.516	0.507	0.445	0.455	0.475	9.4
3+4-Methylphenols		1.631	1.529	1.527	1.375	1.342	1.415	10.9
n-Nitroso-di-n-propylamine	1.062	1.068	1.011	1.009	0.906	0.915	0.966	8.4
Nitrobenzene-d5		0.400	0.390	0.394	0.365	0.370	0.377	4.8
Hexachloroethane		0.610	0.591	0.595	0.548	0.551	0.565	6.4
Nitrobenzene		0.419	0.410	0.411	0.377	0.384	0.393	5.2
Isophorone		0.692	0.668	0.674	0.617	0.629	0.645	5.1
2-Nitrophenol		0.161	0.173	0.190	0.178	0.184	0.179	5.3
2,4-Dimethylphenol		0.252	0.244	0.250	0.236	0.239	0.241	3.3
bis(2-Chloroethoxy) methane		0.435	0.419	0.430	0.387	0.393	0.405	5.6
2,4-Dichlorophenol		0.296	0.296	0.300	0.277	0.286	0.287	4.3
Naphthalene		1.178	1.126	1.131	0.998	1.023	1.055	8.6
4-Chloroaniline		0.397	0.390	0.386	0.345	0.349	0.365	7.1
Hexachlorobutadiene		0.213	0.207	0.208	0.188	0.194	0.197	6.2
Caprolactam		0.096	0.092	0.099	0.091	0.094	0.094	3.0
4-Chloro-3-methylphenol		0.331	0.322	0.327	0.302	0.313	0.313	4.4
2-Methylnaphthalene		0.751	0.719	0.710	0.638	0.654	0.672	8.3
Hexachlorocyclopentadiene		0.170	0.181	0.201	0.200	0.189	0.188	5.7
2,4,6-Trichlorophenol		0.399	0.384	0.412	0.380	0.382	0.387	3.4
2-Fluorobiphenyl		1.582	1.459	1.408	1.225	1.224	1.310	13.4
2,4,5-Trichlorophenol		0.429	0.419	0.418	0.401	0.411	0.409	4.1
1,1-Biphenyl		1.761	1.671	1.654	1.490	1.491	1.553	9.3
2-Chloronaphthalene		1.345	1.288	1.273	1.164	1.174	1.210	7.8
2-Nitroaniline		0.360	0.358	0.365	0.351	0.365	0.359	2.0
Dimethylphthalate		1.447	1.406	1.404	1.299	1.318	1.343	5.8
Acenaphthylene		1.915	1.820	1.830	1.660	1.678	1.728	7.5

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: WEST04
Lab Code: CHEM Case No.: Q1109 SAS No.: Q1109 SDG No.: Q1109
Instrument ID: BNA_F Calibration Date(s): 01/10/2025 01/10/2025
Calibration Time(s): 11:58 17:45

LAB FILE ID:		RRF2.5 = BF141109.D			RRF005 = BF141110.D		RRF010 = BF141111.D	
		RRF020 = BF141112.D			RRF040 = BF141113.D		RRF050 = BF141117.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.308	0.317	0.328	0.312	0.315	0.312	3.3
3-Nitroaniline		0.321	0.324	0.333	0.316	0.320	0.317	3.7
Acenaphthene		1.317	1.267	1.274	1.165	1.185	1.208	6.7
2,4-Dinitrophenol			0.097	0.133	0.155	0.168	0.148	19.4
4-Nitrophenol		0.224	0.241	0.255	0.250	0.264	0.248	5.3
Dibenzofuran		1.866	1.736	1.738	1.586	1.585	1.642	8.8
2,4-Dinitrotoluene		0.372	0.401	0.423	0.395	0.424	0.402	4.7
Diethylphthalate		1.416	1.397	1.378	1.253	1.289	1.311	6.6
4-Chlorophenyl-phenylether		0.709	0.672	0.663	0.594	0.593	0.622	9.7
Fluorene		1.488	1.403	1.359	1.197	1.213	1.276	11.1
4-Nitroaniline		0.314	0.312	0.315	0.300	0.323	0.311	3.1
4,6-Dinitro-2-methylphenol		0.077	0.095	0.120	0.129	0.130	0.116	18.4
n-Nitrosodiphenylamine		0.706	0.677	0.683	0.651	0.603	0.645	7.4
2,4,6-Tribromophenol		0.233	0.233	0.235	0.219	0.230	0.228	3.0
4-Bromophenyl-phenylether		0.241	0.237	0.238	0.232	0.220	0.230	4.2
Hexachlorobenzene		0.273	0.263	0.266	0.257	0.244	0.256	4.8
Atrazine		0.213	0.201	0.170	0.141	0.184	0.174	16.3
Pentachlorophenol		0.142	0.162	0.171	0.173	0.169	0.164	6.6
Phenanthrene		1.218	1.152	1.122	1.064	1.014	1.074	9.0
Anthracene		1.162	1.111	1.095	1.042	0.994	1.044	8.2
Carbazole		1.052	1.025	1.011	0.944	0.935	0.957	8.1
Di-n-butylphthalate		1.186	1.177	1.139	1.087	1.079	1.096	7.3
Fluoranthene		1.211	1.188	1.104	1.037	1.051	1.071	9.7
Pyrene		1.952	1.977	2.081	1.848	1.839	1.910	6.1
Terphenyl-d14		1.444	1.424	1.483	1.272	1.280	1.346	8.4
Butylbenzylphthalate		0.583	0.629	0.663	0.614	0.677	0.637	5.4
3,3-Dichlorobenzidine		0.434	0.422	0.444	0.418	0.426	0.433	2.7
Benzo (a) anthracene		1.473	1.420	1.427	1.245	1.329	1.361	6.3
Chrysene		1.382	1.316	1.336	1.207	1.237	1.274	5.5
Bis (2-ethylhexyl) phthalate		0.734	0.768	0.807	0.726	0.819	0.764	5.4
Di-n-octyl phthalate		0.950	1.039	1.169	1.151	1.235	1.154	10.5
Benzo (b) fluoranthene		1.461	1.415	1.290	1.201	1.315	1.290	9.0
Benzo (k) fluoranthene		1.128	1.102	1.141	1.043	1.038	1.090	3.7
Benzo (a) pyrene		1.103	1.090	1.095	1.039	1.048	1.064	3.3
Indeno (1,2,3-cd) pyrene		1.472	1.420	1.558	1.472	1.421	1.465	4.5

All other compounds must meet a minimum RRF of 0.010.

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: WEST04
Lab Code: CHEM Case No.: Q1109 SAS No.: Q1109 SDG No.: Q1109
Instrument ID: BNA_F Calibration Date(s): 01/10/2025 01/10/2025
Calibration Time(s): 11:58 17:45

LAB FILE ID:		RRF2.5 = BF141109.D		RRF005 = BF141110.D		RRF010 = BF141111.D		
		RRF020 = BF141112.D		RRF040 = BF141113.D		RRF050 = BF141117.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo(a,h)anthracene		1.174	1.178	1.262	1.197	1.150	1.185	5.0
Benzo(g,h,i)perylene		1.256	1.221	1.347	1.263	1.195	1.233	7.6
1,2,4,5-Tetrachlorobenzene		0.632	0.601	0.604	0.554	0.552	0.574	6.8
1,4-Dioxane		0.616	0.596	0.598	0.563	0.573	0.577	5.2
2,3,4,6-Tetrachlorophenol		0.353	0.358	0.354	0.332	0.333	0.340	4.7