

Surrogate Summary

SDG No.: Q1109

Client: Weston Solutions

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1109-02	TAPIAL1-MW04S-011525-00-T2	1,2-Dichloroethane-d4	50	55.1	110	81	118
		Dibromofluoromethane	50	53.0	106	80	119
		Toluene-d8	50	50.1	100	89	112
		4-Bromofluorobenzene	50	47.6	95	85	114
Q1109-04	TAP-TB-01-011525	1,2-Dichloroethane-d4	50	56.9	114	81	118
		Dibromofluoromethane	50	53.0	106	80	119
		Toluene-d8	50	49.7	99	89	112
		4-Bromofluorobenzene	50	46.9	94	85	114

Surrogate Summary

SDG No.: Q1109

Client: Weston Solutions

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VN0116WBL01	VN0116WBL01	1,2-Dichloroethane-d4	50	57.4	115	81	118
		Dibromofluoromethane	50	53.1	106	80	119
		Toluene-d8	50	50.2	100	89	112
		4-Bromofluorobenzene	50	47.9	96	85	114
VN0116WBS02	VN0116WBS02	1,2-Dichloroethane-d4	50	52.5	105	81	118
		Dibromofluoromethane	50	53.9	108	80	119
		Toluene-d8	50	56.1	112	89	112
		4-Bromofluorobenzene	50	55.9	112	85	114
VN0116WBSD0	VN0116WBSD02	1,2-Dichloroethane-d4	50	54.2	108	81	118
		Dibromofluoromethane	50	53.3	107	80	119
		Toluene-d8	50	55.3	111	89	112
		4-Bromofluorobenzene	50	54.2	108	85	114