

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WEST04
 Lab Code: CHEM Case No.: Q1109 SAS No.: Q1109 SDG NO.: Q1109
 Lab File ID: VN085465.D Date Analyzed: 01/16/2025
 Instrument ID: MSVOA_N Time Analyzed: 12:30
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	154064	13.788				
UPPER LIMIT	308128	14.288				
LOWER LIMIT	77032	13.288				
EPA SAMPLE NO.						
TAPIAL1-MW04S-011525-00-T2	106916	13.79				
TAP-TB-01-011525	107148	13.79				
VN0116WBL01	123871	13.79				
VN0116WBS02	152840	13.79				
VN0116WBSD02	135710	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.