

Preparation Blank Summary

Client: Major Products

SDG No.: Q1111

Project: Floor Drain Water 2025

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB134301BL BOD5	mg/L	< 0.2000	0.2000	U	0.17	2.0	01/16/2025
Sample ID: LB134342BL TSS	mg/L	< 2.0000	2.0000	U	1	4	01/17/2025