

Laboratory Control Sample Summary

Client: Major Products

SDG No.: Q1111

Project: Floor Drain Water 2025

Run No.: LB134301

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134301BS							
BOD5	mg/L	198	193		97	1	84.6-115.4	01/16/2025

Laboratory Control Sample Summary

Client: Major Products

SDG No.: Q1111

Project: Floor Drain Water 2025

Run No.: LB134342

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134342BS							
TSS	mg/L	550	520		94	1	90-110	01/17/2025