

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1113

Project: Linden Generating Station - LRSA

RunNo.: LB134325

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	01/17/2025
Sample ID: CCV1 Ammonia as N	mg/L	1	1	100	90-110	01/17/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	01/17/2025
Sample ID: CCV3 Ammonia as N	mg/L	1.1	1	110	90-110	01/17/2025
Sample ID: CCV4 Ammonia as N	mg/L	1	1	100	90-110	01/17/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1113

Project: Linden Generating Station - LRSA

RunNo.: LB134365

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	49.329	50	99	95-105	01/22/2025
Sample ID: CCV1 COD	mg/L	50.319	50	101	95-105	01/22/2025
Sample ID: CCV2 COD	mg/L	49.329	50	99	95-105	01/22/2025
Sample ID: CCV3 COD	mg/L	49.329	50	99	95-105	01/22/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1113

Project: Linden Generating Station - LRSA

RunNo.: LB134376

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
Field pH		pH	7	7	100	90-110	01/16/2025
Sample ID:	CCV1						
Field pH		pH	7	7	100	90-110	01/16/2025
Sample ID:	CCV2						
Field pH		pH	7.01	7	100	90-110	01/16/2025

Initial and Continuing Calibration Verification

Client: Parkway Generation Operating LLC

SDG No.: Q1113

Project: Linden Generating Station - LRSA

RunNo.: LB134376

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------