

Matrix Spike Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q1113
Project:	Linden Generating Station - LRSA	Sample ID:	Q1113-01
Client ID:	DSN002MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/L	75-125	4.30	OR	3.40	OR	1	1	90		01/17/2025
COD	mg/L	75-125	102		54.3		50.0	1	95		01/22/2025

Matrix Spike Summary

Client:	Parkway Generation Operating LLC	SDG No.:	Q1113
Project:	Linden Generating Station - LRSA	Sample ID:	Q1113-01
Client ID:	DSN002MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Ammonia as N	mg/L	75-125	4.10	OR	3.40	OR	1	1	70	*	01/17/2025
COD	mg/L	75-125	101		54.3		50.0	1	93		01/22/2025