

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Parkway Generation Operating LLC **SDG No.:** Q1113
Contract: PSEG04 **Lab Code:** CHEM **Case No.:** Q1113 **SAS No.:** Q1113
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Chromium	531	520	102	95 - 105	P	01/20/2025	10:47	LB134350
	Zinc	1010	1000	101	95 - 105	P	01/20/2025	10:47	LB134350

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Parkway Generation Operating LLC **SDG No.:** Q1113
Contract: PSEG04 **Lab Code:** CHEM **Case No.:** Q1113 **SAS No.:** Q1113
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Chromium	9.71	10.0	97	80 - 120	P	01/20/2025	10:53	LB134350
	Zinc	43.5	40.0	109	80 - 120	P	01/20/2025	10:53	LB134350

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Parkway Generation Operating LLC **SDG No.:** Q1113
Contract: PSEG04 **Lab Code:** CHEM **Case No.:** Q1113 **SAS No.:** Q1113
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Chromium	1000	1000	100	90 - 110	P	01/20/2025	11:31	LB134350
	Zinc	2480	2500	99	90 - 110	P	01/20/2025	11:31	LB134350
CCV02	Chromium	995	1000	100	90 - 110	P	01/20/2025	12:29	LB134350
	Zinc	2470	2500	99	90 - 110	P	01/20/2025	12:29	LB134350
CCV03	Chromium	976	1000	98	90 - 110	P	01/20/2025	13:35	LB134350
	Zinc	2410	2500	96	90 - 110	P	01/20/2025	13:35	LB134350
CCV04	Chromium	991	1000	99	90 - 110	P	01/20/2025	14:38	LB134350
	Zinc	2470	2500	99	90 - 110	P	01/20/2025	14:38	LB134350
CCV05	Chromium	987	1000	99	90 - 110	P	01/20/2025	15:37	LB134350
	Zinc	2490	2500	100	90 - 110	P	01/20/2025	15:37	LB134350
CCV06	Chromium	966	1000	97	90 - 110	P	01/20/2025	16:47	LB134350
	Zinc	2410	2500	96	90 - 110	P	01/20/2025	16:47	LB134350
CCV07	Chromium	987	1000	99	90 - 110	P	01/20/2025	17:33	LB134350
	Zinc	2450	2500	98	90 - 110	P	01/20/2025	17:33	LB134350
CCV08	Chromium	969	1000	97	90 - 110	P	01/20/2025	17:52	LB134350
	Zinc	2380	2500	95	90 - 110	P	01/20/2025	17:52	LB134350