

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Parkway Generation Operating LLC **SDG No.:** Q1113
Contract: PSEG04 **Lab Code:** CHEM **Case No.:** Q1113 **SAS No.:** Q1113
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Chromium	57.5	52.0	111	42	62	01/20/2025	11:10	LB134350
	Zinc	3.10			-40	40	01/20/2025	11:10	LB134350
ICSAB01	Chromium	579	542	107	460	624	01/20/2025	11:19	LB134350
	Zinc	1070	952	112	809	1095	01/20/2025	11:19	LB134350
ICSA	Chromium	48.2	52.0	93	42	62	01/20/2025	11:23	LB134350
	Zinc	-9.98			-40	40	01/20/2025	11:23	LB134350
ICSAB	Chromium	576	542	106	460	624	01/20/2025	11:27	LB134350
	Zinc	1040	952	109	809	1095	01/20/2025	11:27	LB134350