

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Parkway Generation Operating LLC **level:** low **sdg no.:** Q1113
contract: PSEG04 **lab code:** CHEM **case no.:** Q1113 **sas no.:** Q1113
matrix: Water **sample id:** Q1113-01 **client id:** DSN002MS
Percent Solids for Sample: NA **Spiked ID:** Q1113-01MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	75 - 125	215		2.96	J	200	106		P
Zinc	ug/L	75 - 125	287		179		100	108		P

metals

- 5a -

MATRIX SPIKE DUPLICATE SUMMARY

client: Parkway Generation Operating LLC **level:** low **sdg no.:** Q1113
contract: PSEG04 **lab code:** CHEM **case no.:** Q1113 **sas no.:** Q1113
matrix: Water **sample id:** Q1113-01 **client id:** DSN002MSD
Percent Solids for Sample: NA **Spiked ID:** Q1113-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	75 - 125	214		2.96	J	200	105		P
Zinc	ug/L	75 - 125	285		179		100	106		P