

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Parkway Generation Operating LLC

SDG No.: Q1113

Contract: PSEG04

Lab Code: CHEM

Case No.: Q1113

SAS No.: Q1113

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166102BS							
Chromium	ug/L	200	196		98	85 - 115	P
Zinc	ug/L	100	99.2		99	85 - 115	P